

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051122\
 Data File : VY008704.D
 Acq On : 11 May 2022 14:06
 Operator : KP/MD
 Sample : N2760-01
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-34

Quant Time: May 12 06:28:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 09 19:21:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	125897	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	202270	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	186585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	83591	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	58884	54.585	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.180%
35) Dibromofluoromethane	7.716	113	65017	52.161	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.320%
50) Toluene-d8	10.179	98	234622	53.414	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.820%
62) 4-Bromofluorobenzene	12.483	95	71455	45.216	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051122\
Data File : VY008704.D
Acq On : 11 May 2022 14:06
Operator : KP/MD
Sample : N2760-01
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-34

Quant Time: May 12 06:28:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
Quant Title : SW846 8260
QLast Update : Mon May 09 19:21:20 2022
Response via : Initial Calibration

