

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051122\
 Data File : VY008715.D
 Acq On : 11 May 2022 18:24
 Operator : KP/MD
 Sample : N2781-05
 Misc : 5.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P003-SS003-2430-01

Quant Time: May 12 06:31:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 09 19:21:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	205764	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	330613	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	306132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	141232	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	100498	57.001	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.000%
35) Dibromofluoromethane	7.716	113	105629	51.845	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.700%
50) Toluene-d8	10.179	98	384428	53.544	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.080%
62) 4-Bromofluorobenzene	12.483	95	120971	46.833	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051122\
Data File : VY008715.D
Acq On : 11 May 2022 18:24
Operator : KP/MD
Sample : N2781-05
Misc : 5.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P003-SS003-2430-01

Quant Time: May 12 06:31:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
Quant Title : SW846 8260
QLast Update : Mon May 09 19:21:20 2022
Response via : Initial Calibration

