

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051122\
 Data File : VY008730.D
 Acq On : 12 May 2022 01:21
 Operator : KP/MD
 Sample : N2781-13
 Misc : 5.44g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P011-SS003-1824-01

Quant Time: May 12 09:29:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 09 19:21:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	95631	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	174237	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	159099	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	68729	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	62457	76.221	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	152.440%
35) Dibromofluoromethane	7.716	113	56251	52.389	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.780%
50) Toluene-d8	10.179	98	210556	55.647	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.300%
62) 4-Bromofluorobenzene	12.483	95	62090	45.612	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051122\
Data File : VY008730.D
Acq On : 12 May 2022 01:21
Operator : KP/MD
Sample : N2781-13
Misc : 5.44g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P011-SS003-1824-01

Quant Time: May 12 09:29:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
Quant Title : SW846 8260
QLast Update : Mon May 09 19:21:20 2022
Response via : Initial Calibration

