

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051122\
 Data File : VY008704.D
 Acq On : 11 May 2022 14:06
 Operator : KP/MD
 Sample : N2760-01
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-34

Quant Time: May 12 06:28:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 09 19:21:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	125897	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	202270	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	186585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	83591	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	58884	54.585	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.180%
35) Dibromofluoromethane	7.716	113	65017	52.161	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.320%
50) Toluene-d8	10.179	98	234622	53.414	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.820%
62) 4-Bromofluorobenzene	12.483	95	71455	45.216	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.440%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051122\
Data File : VY008704.D
Acq On : 11 May 2022 14:06
Operator : KP/MD
Sample : N2760-01
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-34

Quant Time: May 12 06:28:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
Quant Title : SW846 8260
QLast Update : Mon May 09 19:21:20 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008704.D
Acq: 11 May 2022 14:06




Instrument :
MSVOA_Y
ClientSampleId :
TP-34


Tgt Ion:168 Resp: 125897
Ion Ratio Lower Upper
168 100
99 45.8 46.9 70.3#



#33
1,2-Dichloroethane-d4
Concen: 54.585 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY008704.D
Acq: 11 May 2022 14:06

Tgt Ion: 65 Resp: 58884
Ion Ratio Lower Upper
65 100
67 54.6 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY008704.D

Acq: 11 May 2022 14:06

Instrument :

MSVOA_Y

ClientSampleId :

TP-34





Tgt Ion:114 Resp: 202270

Ion Ratio Lower Upper

114 100

63 16.3 0.0 43.2

88 13.5 0.0 31.8



#35

Dibromofluoromethane

Concen: 52.161 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY008704.D

Acq: 11 May 2022 14:06




Tgt Ion:113 Resp: 65017


Ion Ratio Lower Upper

113 100

111 104.4 82.6 124.0

192 20.8 17.6 26.4









#50
Toluene-d8
Concen: 53.414 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008704.D
Acq: 11 May 2022 14:06




Instrument :
MSVOA_Y
ClientSampleId :
TP-34


Tgt Ion: 98 Resp: 234622
Ion Ratio Lower Upper
98 100
100 65.1 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 45.216 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008704.D
Acq: 11 May 2022 14:06

Tgt Ion: 95 Resp: 71455
Ion Ratio Lower Upper
95 100
174 94.9 0.0 177.0
176 92.6 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 117
Delta R.T. 0.001 min
Lab File: VY008704.D
Acq: 11 May 2022 14:06

Instrument :
MSVOA_Y
ClientSampleId :
TP-34

Tgt Ion:117 Resp: 186585
Ion Ratio Lower Upper
117 100
82 52.1 43.0 64.4
119 32.6 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008704.D
Acq: 11 May 2022 14:06

Tgt Ion:152 Resp: 83591
Ion Ratio Lower Upper
152 100
115 53.4 28.2 84.7
150 155.6 0.0 347.6

