

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051122\
 Data File : VY008711.D
 Acq On : 11 May 2022 16:50
 Operator : KP/MD
 Sample : N2781-01
 Misc : 5.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P003-SS003-0006-01

Quant Time: May 12 06:30:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 09 19:21:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	205877	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	333805	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	302551	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	138672	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	95266	54.004	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.000%
35) Dibromofluoromethane	7.716	113	105501	51.287	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.580%
50) Toluene-d8	10.179	98	383612	52.920	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.840%
62) 4-Bromofluorobenzene	12.483	95	121758	46.687	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.380%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051122\
Data File : VY008711.D
Acq On : 11 May 2022 16:50
Operator : KP/MD
Sample : N2781-01
Misc : 5.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P003-SS003-0006-01

Quant Time: May 12 06:30:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
Quant Title : SW846 8260
QLast Update : Mon May 09 19:21:20 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008711.D
Acq: 11 May 2022 16:50




Instrument :
MSVOA_Y
ClientSampleId :
P003-SS003-0006-01


Tgt Ion:168 Resp: 205877
Ion Ratio Lower Upper
168 100
99 46.7 46.9 70.3#



#33
1,2-Dichloroethane-d4
Concen: 54.004 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY008711.D
Acq: 11 May 2022 16:50

Tgt Ion: 65 Resp: 95266
Ion Ratio Lower Upper
65 100
67 57.2 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1127

Delta R.T. -0.006 min

Lab File: VY008711.D

Acq: 11 May 2022 16:50

Instrument :

MSVOA_Y

ClientSampleId :

P003-SS003-0006-01





Tgt Ion:114 Resp: 333805

Ion Ratio Lower Upper

114 100

63 16.6 0.0 43.2

88 13.1 0.0 31.8



#35

Dibromofluoromethane

Concen: 51.287 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY008711.D

Acq: 11 May 2022 16:50




Tgt Ion:113 Resp: 105501


Ion Ratio Lower Upper

113 100

111 101.2 82.6 124.0

192 20.9 17.6 26.4









#50
Toluene-d8
Concen: 52.920 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY008711.D
Acq: 11 May 2022 16:50




Instrument :
MSVOA_Y
ClientSampleId :
P003-SS003-0006-01


Tgt Ion: 98 Resp: 383612
Ion Ratio Lower Upper
98 100
100 64.9 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 46.687 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008711.D
Acq: 11 May 2022 16:50

Tgt Ion: 95 Resp: 121758
Ion Ratio Lower Upper
95 100
174 95.1 0.0 177.0
176 92.6 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 117
Delta R.T. 0.000 min
Lab File: VY008711.D
Acq: 11 May 2022 16:50

Instrument :
MSVOA_Y
ClientSampleId :
P003-SS003-0006-01

Tgt Ion:	117	Resp:	302551
Ion	Ratio	Lower	Upper
117	100		
82	49.3	43.0	64.4
119	30.8	25.4	38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008711.D
Acq: 11 May 2022 16:50

Tgt Ion:	152	Resp:	138672
Ion	Ratio	Lower	Upper
152	100		
115	54.0	28.2	84.7
150	155.9	0.0	347.6

