


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051122\
Data File : VY008723.D
Acq On : 11 May 2022 21:30
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
Title : SW846 8260
Last Update : Mon May 09 19:21:20 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.2	9470	PASS
75	95	30	60	50.1	29214	PASS
95	95	100	100	100.0	58301	PASS
96	95	5	9	6.5	3787	PASS
173	174	0.00	2	1.1	616	PASS
174	95	50	100	99.0	57693	PASS
175	174	5	9	7.1	4074	PASS
176	174	95	101	96.7	55803	PASS
177	176	5	9	6.8	3798	PASS