

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051225\
 Data File : VY022216.D
 Acq On : 12 May 2025 12:02
 Operator : SY/MD
 Sample : Q1983-14
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-636-VOC-03

Quant Time: May 13 02:18:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	142531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	248448	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	213483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	82030	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	80587	61.212	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.420%
35) Dibromofluoromethane	7.634	113	93645	57.053	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.100%
50) Toluene-d8	10.109	98	304121	49.147	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.300%
62) 4-Bromofluorobenzene	12.408	95	75859	36.416	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	72.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051225\
Data File : VY022216.D
Acq On : 12 May 2025 12:02
Operator : SY/MD
Sample : Q1983-14
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-636-VOC-03

Quant Time: May 13 02:18:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

