

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051225\
 Data File : VY022208.D
 Acq On : 12 May 2025 08:26
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 13 02:13:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	72	0.00
2 T	Dichlorodifluoromethane	0.426	0.369	13.4	68	0.00
3 P	Chloromethane	0.607	0.540	11.0	65	0.00
4 C	Vinyl Chloride	0.745	0.725	2.7#	72	0.00
5 T	Bromomethane	0.642	0.556	13.4	68	0.00
6 T	Chloroethane	0.508	0.512	-0.8	75	0.00
7 T	Trichlorofluoromethane	0.983	0.928	5.6	70	0.00
8 T	Diethyl Ether	0.256	0.234	8.6	68	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.533	0.536	-0.6	79	0.00
10 T	Methyl Iodide	0.617	0.617	0.0	71	0.00
11 T	Tert butyl alcohol	0.025	0.022	12.0	62	-0.02
12 CM	1,1-Dichloroethene	0.497	0.472	5.0#	73	0.00
13 T	Acrolein	0.046	0.010	78.3#	17#	0.00
14 T	Allyl chloride	0.579	0.594	-2.6	77	0.00
15 T	Acrylonitrile	0.094	0.089	5.3	69	0.00
16 T	Acetone	0.070	0.064	8.6	74	0.00
17 T	Carbon Disulfide	1.585	1.449	8.6	70	0.00
18 T	Methyl Acetate	0.216	0.248	-14.8	81	0.00
19 T	Methyl tert-butyl Ether	1.113	1.089	2.2	70	0.00
20 T	Methylene Chloride	0.584	0.542	7.2	74	0.00
21 T	trans-1,2-Dichloroethene	0.557	0.550	1.3	75	0.01
22 T	Diisopropyl ether	1.288	1.384	-7.5	78	0.00
23 T	Vinyl Acetate	0.694	0.708	-2.0	72	0.00
24 P	1,1-Dichloroethane	0.893	0.921	-3.1	78	0.00
25 T	2-Butanone	0.105	0.102	2.9	70	0.00
26 T	2,2-Dichloropropane	0.782	0.839	-7.3	82	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.625	-0.5	74	0.00
28 T	Bromochloromethane	0.347	0.354	-2.0	75	0.00
29 T	Tetrahydrofuran	0.068	0.068	0.0	68	0.00
30 C	Chloroform	0.990	1.025	-3.5#	78	0.00
31 T	Cyclohexane	0.763	0.768	-0.7	79	0.00
32 T	1,1,1-Trichloroethane	0.896	0.927	-3.5	79	0.00
33 S	1,2-Dichloroethane-d4	0.462	0.442	4.3	69	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	69	0.00
35 S	Dibromofluoromethane	0.330	0.337	-2.1	70	0.00
36 T	1,1-Dichloropropene	0.444	0.486	-9.5	78	0.00
37 T	Ethyl Acetate	0.159	0.159	0.0	68	0.00
38 T	Carbon Tetrachloride	0.530	0.587	-10.8	79	0.00
39 T	Methylcyclohexane	0.558	0.605	-8.4	76	0.00
40 TM	Benzene	1.387	1.497	-7.9	77	0.00
41 T	Methacrylonitrile	0.087	0.082	5.7	61	0.00
42 TM	1,2-Dichloroethane	0.343	0.372	-8.5	76	0.00
43 T	Isopropyl Acetate	0.306	0.307	-0.3	68	0.00
44 TM	Trichloroethene	0.384	0.403	-4.9	75	0.00
45 C	1,2-Dichloropropane	0.307	0.343	-11.7#	79	0.00
46 T	Dibromomethane	0.192	0.202	-5.2	74	0.00
47 T	Bromodichloromethane	0.480	0.533	-11.0	78	0.00
48 T	Methyl methacrylate	0.141	0.154	-9.2	70	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051225\
 Data File : VY022208.D
 Acq On : 12 May 2025 08:26
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 13 02:13:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	68	0.00
50 S	Toluene-d8	1.245	1.296	-4.1	70	0.00
51 T	4-Methyl-2-Pentanone	0.160	0.175	-9.4	70	0.00
52 CM	Toluene	0.898	1.011	-12.6#	78	0.00
53 T	t-1,3-Dichloropropene	0.411	0.443	-7.8	73	0.00
54 T	cis-1,3-Dichloropropene	0.489	0.535	-9.4	76	0.00
55 T	1,1,2-Trichloroethane	0.256	0.273	-6.6	73	0.00
56 T	Ethyl methacrylate	0.293	0.307	-4.8	67	0.00
57 T	1,3-Dichloropropane	0.411	0.455	-10.7	76	0.00
58 T	2-Chloroethyl Vinyl ether	0.143	0.156	-9.1	67	0.00
59 T	2-Hexanone	0.106	0.114	-7.5	68	0.00
60 T	Dibromochloromethane	0.355	0.383	-7.9	75	0.00
61 T	1,2-Dibromoethane	0.242	0.257	-6.2	73	0.00
62 S	4-Bromofluorobenzene	0.419	0.438	-4.5	71	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	70	0.00
64 T	Tetrachloroethene	0.472	0.488	-3.4	75	0.00
65 PM	Chlorobenzene	1.118	1.194	-6.8	77	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.428	-8.6	77	0.00
67 C	Ethyl Benzene	1.829	2.027	-10.8#	77	0.00
68 T	m/p-Xylenes	0.728	0.816	-12.1	77	0.00
69 T	o-Xylene	0.671	0.747	-11.3	76	0.00
70 T	Styrene	1.126	1.288	-14.4	77	0.00
71 P	Bromoform	0.229	0.241	-5.2	73	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
73 T	Isopropylbenzene	3.341	3.598	-7.7	78	0.00
74 T	N-amyl acetate	0.574	0.562	2.1	68	0.00
75 P	1,1,2,2-Tetrachloroethane	0.563	0.567	-0.7	74	0.00
76 T	1,2,3-Trichloropropane	0.428	0.440	-2.8	78	0.00
77 T	Bromobenzene	0.844	0.861	-2.0	75	0.00
78 T	n-propylbenzene	3.978	4.341	-9.1	78	0.00
79 T	2-Chlorotoluene	2.314	2.470	-6.7	78	0.00
80 T	1,3,5-Trimethylbenzene	2.752	3.019	-9.7	77	0.00
81 T	trans-1,4-Dichloro-2-butene	0.178	0.182	-2.2	74	0.00
82 T	4-Chlorotoluene	2.405	2.580	-7.3	78	0.00
83 T	tert-Butylbenzene	2.468	2.677	-8.5	77	0.00
84 T	1,2,4-Trimethylbenzene	2.753	2.998	-8.9	77	0.00
85 T	sec-Butylbenzene	3.612	3.983	-10.3	79	0.00
86 T	p-Isopropyltoluene	3.067	3.379	-10.2	78	0.00
87 T	1,3-Dichlorobenzene	1.714	1.796	-4.8	78	0.00
88 T	1,4-Dichlorobenzene	1.698	1.748	-2.9	78	0.00
89 T	n-Butylbenzene	2.742	2.989	-9.0	78	0.00
90 T	Hexachloroethane	0.674	0.710	-5.3	80	0.00
91 T	1,2-Dichlorobenzene	1.495	1.544	-3.3	76	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.090	0.084	6.7	72	0.00
93 T	1,2,4-Trichlorobenzene	0.847	0.832	1.8	73	0.00
94 T	Hexachlorobutadiene	0.521	0.526	-1.0	79	0.00
95 T	Naphthalene	1.408	1.323	6.0	66	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051225\
Data File : VY022208.D
Acq On : 12 May 2025 08:26
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: May 13 02:13:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.731	0.724	1.0	74	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6