

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051225\
 Data File : VY022221.D
 Acq On : 12 May 2025 13:55
 Operator : SY/MD
 Sample : Q1983-44
 Misc : 4.54g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-636-VOC-08

Quant Time: May 13 02:20:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	142954	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	259399	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	232592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	92203	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	91564	69.344	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.680%
35) Dibromofluoromethane	7.634	113	98348	57.388	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.780%
50) Toluene-d8	10.109	98	323694	50.102	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.200%
62) 4-Bromofluorobenzene	12.408	95	86607	39.820	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.640%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051225\
Data File : VY02221.D
Acq On : 12 May 2025 13:55
Operator : SY/MD
Sample : Q1983-44
Misc : 4.54g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-636-VOC-08

Quant Time: May 13 02:20:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY022221.D
Acq: 12 May 2025 13:55




Instrument :
MSVOA_Y
ClientSampleId :
OR-636-VOC-08


Tgt Ion:168 Resp: 142954
Ion Ratio Lower Upper
168 100
99 49.2 43.1 64.7



#33
1,2-Dichloroethane-d4
Concen: 69.344 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VY022221.D
Acq: 12 May 2025 13:55

Tgt Ion: 65 Resp: 91564
Ion Ratio Lower Upper
65 100
67 52.1 0.0 105.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.616 min Scan# 1131
Delta R.T. 0.000 min
Lab File: VY022221.D
Acq: 12 May 2025 13:55




Instrument :
MSVOA_Y
ClientSampleId :
OR-636-VOC-08


Tgt Ion:114 Resp: 259399
Ion Ratio Lower Upper
114 100
63 17.9 0.0 35.4
88 14.6 0.0 28.2



#35
Dibromofluoromethane
Concen: 57.388 ug/l
RT: 7.634 min Scan# 970
Delta R.T. 0.000 min
Lab File: VY022221.D
Acq: 12 May 2025 13:55

Tgt Ion:113 Resp: 98348
Ion Ratio Lower Upper
113 100
111 101.3 81.8 122.8
192 19.4 15.6 23.4









#50
Toluene-d8
Concen: 50.102 ug/l
RT: 10.109 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY022221.D
Acq: 12 May 2025 13:55




Instrument :
MSVOA_Y
ClientSampleId :
OR-636-VOC-08


Tgt Ion: 98 Resp: 323694
Ion Ratio Lower Upper
98 100
100 63.1 51.6 77.4



#62
4-Bromofluorobenzene
Concen: 39.820 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY022221.D
Acq: 12 May 2025 13:55

Tgt Ion: 95 Resp: 86607
Ion Ratio Lower Upper
95 100
174 87.9 0.0 179.0
176 85.0 0.0 171.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY022221.D

Acq: 12 May 2025 13:55

Instrument :

MSVOA_Y

ClientSampleId :

OR-636-VOC-08





Tgt Ion:117 Resp: 232592

Ion Ratio Lower Upper

117 100

82 52.1 42.1 63.1

119 30.9 25.7 38.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.347 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VY022221.D

Acq: 12 May 2025 13:55

Tgt Ion:152 Resp: 92203

Ion Ratio Lower Upper

152 100

115 55.4 28.0 84.0

150 159.5 0.0 345.6

