

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051325\
 Data File : VY022237.D
 Acq On : 13 May 2025 12:42
 Operator : SY/MD
 Sample : Q2002-01RE
 Misc : 5.27g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-02-05092025RE

Quant Time: May 13 12:55:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	202993	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	342454	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	234031	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	72423	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	96925	51.693	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 103.380%	
35) Dibromofluoromethane	7.634	113	114279	50.512	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.020%	
50) Toluene-d8	10.103	98	332944	39.035	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 78.080%	
62) 4-Bromofluorobenzene	12.408	95	75742	26.379	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 52.760%	
Target Compounds						
						Qvalue
20) Methylene Chloride	4.610	84	2602	1.097	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051325\
Data File : VY022237.D
Acq On : 13 May 2025 12:42
Operator : SY/MD
Sample : Q2002-01RE
Misc : 5.27g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-02-05092025RE

Quant Time: May 13 12:55:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

