

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051524\
 Data File : VY018266.D
 Acq On : 15 May 2024 10:44
 Operator : SY/MD
 Sample : VY0515SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0515SBL01

Quant Time: May 16 03:04:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	171606	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	316101	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	279510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	100128	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	80183	63.514	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.020%
35) Dibromofluoromethane	7.710	113	92457	55.655	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.300%
50) Toluene-d8	10.179	98	362754	55.413	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.820%
62) 4-Bromofluorobenzene	12.483	95	103801	49.099	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051524\
Data File : VY018266.D
Acq On : 15 May 2024 10:44
Operator : SY/MD
Sample : VY0515SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0515SBL01

Quant Time: May 16 03:04:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

