

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051524\
 Data File : VY018272.D
 Acq On : 15 May 2024 15:03
 Operator : SY/MD
 Sample : P2482-02
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SW-01-229

Quant Time: May 16 03:06:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	143935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	272139	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	235858	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	85530	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	70868	66.927	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	133.860%	
35) Dibromofluoromethane	7.716	113	79815	55.806	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	111.620%	
50) Toluene-d8	10.179	98	307174	54.503	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	109.000%	
62) 4-Bromofluorobenzene	12.483	95	90858	49.920	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	99.840%	
Target Compounds						
20) Methylene Chloride	4.704	84	9029	6.045	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051524\
Data File : VY018272.D
Acq On : 15 May 2024 15:03
Operator : SY/MD
Sample : P2482-02
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SW-01-229

Quant Time: May 16 03:06:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

