

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051524\
 Data File : VY018278.D
 Acq On : 15 May 2024 17:26
 Operator : SY/MD
 Sample : P2474-02
 Misc : 3.18g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-114-SB01

Quant Time: May 16 03:08:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	145117	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	284188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	246327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	87095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	73327	68.685	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 137.380%	
35) Dibromofluoromethane	7.715	113	82778	55.424	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.840%	
50) Toluene-d8	10.178	98	322989	54.879	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 109.760%	
62) 4-Bromofluorobenzene	12.483	95	93359	49.119	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 98.240%	
Target Compounds						
20) Methylene Chloride	4.704	84	9111	6.050	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051524\
Data File : VY018278.D
Acq On : 15 May 2024 17:26
Operator : SY/MD
Sample : P2474-02
Misc : 3.18g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-114-SB01

Quant Time: May 16 03:08:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

