

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051524\
 Data File : VY018267.D
 Acq On : 15 May 2024 11:20
 Operator : SY/MD
 Sample : P2403-03 2.5 PPB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MDL-SOIL-03-QT2-2024

Quant Time: May 16 03:54:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 03:54:18 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
 Supervised By :Semsettin Yesilyurt 05/16/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	185832	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	337173	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	274967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	114541	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	40856	29.885	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	59.760%		
35) Dibromofluoromethane	7.716	113	47007	26.528	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	53.060%#		
50) Toluene-d8	10.179	98	191225	27.385	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	54.780%#		
62) 4-Bromofluorobenzene	12.483	95	56271	24.954	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	49.900%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.906	85	4073	2.764	ug/l	97
3) Chloromethane	2.113	50	6885	3.490	ug/l	92
4) Vinyl Chloride	2.247	62	7589	3.173	ug/l	99
5) Bromomethane	2.649	94	6068	3.654	ug/l	92
6) Chloroethane	2.790	64	4978	3.280	ug/l	99
7) Trichlorofluoromethane	3.119	101	8191	2.924	ug/l	100
8) Diethyl Ether	3.521	74	3064	3.323	ug/l	93
9) 1,1,2-Trichlorotrifluo...	3.887	101	5231	2.930	ug/l	89
10) Methyl Iodide	4.082	142	4885	2.343	ug/l	97
11) Tert butyl alcohol	4.942	59	2664	19.495	ug/l #	87
12) 1,1-Dichloroethene	3.875	96	5544	3.085	ug/l #	75
13) Acrolein	3.722	56	2655m	18.017	ug/l	
14) Allyl chloride	4.472	41	7066	3.091	ug/l	95
15) Acrylonitrile	5.161	53	5219	14.717	ug/l	92
16) Acetone	3.936	43	4694	17.205	ug/l	91
17) Carbon Disulfide	4.186	76	15201	3.107	ug/l #	94
18) Methyl Acetate	4.460	43	2760	3.551	ug/l	96
19) Methyl tert-butyl Ether	5.216	73	11081	2.768	ug/l	97
20) Methylene Chloride	4.704	84	11851	6.146	ug/l	94
21) trans-1,2-Dichloroethene	5.216	96	5929	3.112	ug/l	91
22) Diisopropyl ether	6.106	45	14389	2.967	ug/l #	92
23) Vinyl Acetate	6.057	43	39319	13.339	ug/l	97
24) 1,1-Dichloroethane	6.003	63	8936	2.991	ug/l	97
25) 2-Butanone	6.978	43	5942	13.625	ug/l	97
26) 2,2-Dichloropropane	6.972	77	8083	2.990	ug/l	85
27) cis-1,2-Dichloroethene	6.984	96	6855	3.045	ug/l	98
28) Bromochloromethane	7.332	49	3846	3.221	ug/l #	94
29) Tetrahydrofuran	7.350	42	4131	14.350	ug/l	94
30) Chloroform	7.496	83	9433	3.025	ug/l	97
31) Cyclohexane	7.783	56	11571	4.080	ug/l #	64
32) 1,1,1-Trichloroethane	7.703	97	7900	2.826	ug/l	97
36) 1,1-Dichloropropene	7.917	75	7579	2.984	ug/l	96
37) Ethyl Acetate	7.069	43	3046	2.849	ug/l #	95
38) Carbon Tetrachloride	7.899	117	7055	2.542	ug/l #	92
39) Methylcyclohexane	9.185	83	9264	2.528	ug/l	97
40) Benzene	8.161	78	22003	2.693	ug/l	92

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051524\
 Data File : VY018267.D
 Acq On : 15 May 2024 11:20
 Operator : SY/MD
 Sample : P2403-03 2.5 PPB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MDL-SOIL-03-QT2-2024

Quant Time: May 16 03:54:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 03:54:18 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
 Supervised By :Semsettin Yesilyurt 05/16/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.313	41	2200	3.563	ug/l #	81
42) 1,2-Dichloroethane	8.240	62	5141	2.892	ug/l	86
43) Isopropyl Acetate	8.270	43	6118	2.792	ug/l	99
44) Trichloroethene	8.941	130	5913	2.658	ug/l	93
45) 1,2-Dichloropropane	9.215	63	4911	2.749	ug/l	91
46) Dibromomethane	9.301	93	2663	2.593	ug/l	99
47) Bromodichloromethane	9.496	83	6330	2.493	ug/l #	97
48) Methyl methacrylate	9.295	41	2405	2.394	ug/l #	94
49) 1,4-Dioxane	9.289	88	960	70.423	ug/l #	96
51) 4-Methyl-2-Pentanone	10.069	43	13412	12.176	ug/l	97
52) Toluene	10.246	92	14015	2.664	ug/l	100
53) t-1,3-Dichloropropene	10.465	75	5840	2.463	ug/l	99
54) cis-1,3-Dichloropropene	9.929	75	7413	2.557	ug/l	95
55) 1,1,2-Trichloroethane	10.654	97	3453	2.448	ug/l #	79
56) Ethyl methacrylate	10.514	69	4207	2.196	ug/l #	74
57) 1,3-Dichloropropane	10.794	76	6001	2.614	ug/l	93
58) 2-Chloroethyl Vinyl ether	9.782	63	10447	11.419	ug/l	95
59) 2-Hexanone	10.837	43	8201	10.648	ug/l	94
60) Dibromochloromethane	10.983	129	4175	2.281	ug/l	95
61) 1,2-Dibromoethane	11.093	107	3302	2.485	ug/l	94
64) Tetrachloroethene	10.721	164	4745	2.592	ug/l	91
65) Chlorobenzene	11.514	112	14373	2.567	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.593	131	4622	2.552	ug/l	91
67) Ethyl Benzene	11.599	91	25434	2.690	ug/l	97
68) m/p-Xylenes	11.709	106	19892	5.320	ug/l	98
69) o-Xylene	12.032	106	8563	2.386	ug/l	89
70) Styrene	12.044	104	14971	2.498	ug/l	100
71) Bromoform	12.209	173	2510	2.504	ug/l #	95
73) Isopropylbenzene	12.331	105	23569	2.744	ug/l	95
74) N-amyl acetate	12.154	43	4485	2.516	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.581	83	3520	2.569	ug/l	92
76) 1,2,3-Trichloropropane	12.642	75	2203m	2.403	ug/l	
77) Bromobenzene	12.611	156	5248	2.660	ug/l	99
78) n-propylbenzene	12.672	91	28649	2.868	ug/l	100
79) 2-Chlorotoluene	12.757	91	15646	2.819	ug/l	97
80) 1,3,5-Trimethylbenzene	12.812	105	18217	2.690	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.379	75	1334	2.803	ug/l	95
82) 4-Chlorotoluene	12.855	91	15591	2.796	ug/l	96
83) tert-Butylbenzene	13.074	119	17530	2.807	ug/l	98
84) 1,2,4-Trimethylbenzene	13.123	105	18396	2.773	ug/l	99
85) sec-Butylbenzene	13.257	105	24523	2.677	ug/l	98
86) p-Isopropyltoluene	13.373	119	20061	2.662	ug/l	99
87) 1,3-Dichlorobenzene	13.367	146	9999	2.624	ug/l	99
88) 1,4-Dichlorobenzene	13.446	146	10032m	2.701	ug/l	
89) n-Butylbenzene	13.702	91	17809	2.611	ug/l	99
90) Hexachloroethane	13.958	117	3674	2.485	ug/l	91
91) 1,2-Dichlorobenzene	13.739	146	8460	2.583	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	14.361	75	416	2.230	ug/l	84
93) 1,2,4-Trichlorobenzene	15.007	180	4315	2.344	ug/l	93
94) Hexachlorobutadiene	15.123	225	2351	2.314	ug/l	95
95) Naphthalene	15.239	128	7120	2.111	ug/l	98
96) 1,2,3-Trichlorobenzene	15.434	180	3409	2.256	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051524\
Data File : VY018267.D
Acq On : 15 May 2024 11:20
Operator : SY/MD
Sample : P2403-03 2.5 PPB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024

Quant Time: May 16 03:54:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Thu May 16 03:54:18 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.789 min Scan# 91

Delta R.T. -0.000 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :

MDL-SOIL-03-QT2-2024

Tgt Ion:168 Resp: 18583

Ion Ratio Lower Upper

168 100





99 48.7 39.1 58.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024

Supervised By :Semsettin Yesilyurt 05/16/2024



#2

Dichlorodifluoromethane

Concen: 2.764 ug/l

RT: 1.906 min Scan# 14

Delta R.T. 0.006 min

Lab File: VY018267.D




Acq: 15 May 2024 11:20


Tgt Ion: 85 Resp: 4073

Ion Ratio Lower Upper

85 100


87 30.4 16.0 48.0





#3
Chloromethane
Concen: 3.490 ug/l
RT: 2.113 min Scan# 41
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024



Tgt Ion: 50 Resp: 688
Ion Ratio Lower Upper
50 100
52 36.6 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#4
Vinyl Chloride
Concen: 3.173 ug/l
RT: 2.247 min Scan# 70
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20


Tgt Ion: 62 Resp: 7589
Ion Ratio Lower Upper
62 100
64 33.3 26.1 39.1





#5
Bromomethane
Concen: 3.654 ug/l
RT: 2.649 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024




Tgt Ion: 94 Resp: 606
Ion Ratio Lower Upper
94 100
96 84.9 74.1 111.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#6
Chloroethane
Concen: 3.280 ug/l
RT: 2.790 min Scan# 159
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Tgt Ion: 64 Resp: 4978
Ion Ratio Lower Upper
64 100
66 31.9 26.1 39.1





#7
Trichlorofluoromethane
Concen: 2.924 ug/l
RT: 3.119 min Scan# 213
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument : MSVOA_Y
ClientSampleId : MDL-SOIL-03-QT2-2024




Tgt Ion: 101 Resp: 819
Ion Ratio Lower Upper
101 100
103 61.7 49.3 73.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#8
Diethyl Ether
Concen: 3.323 ug/l
RT: 3.521 min Scan# 279
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20



Tgt Ion: 74 Resp: 3064
Ion Ratio Lower Upper
74 100
45 76.4 41.3 124.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 2.930 ug/l
RT: 3.887 min Scan# 339
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024

Tgt Ion:101 Resp: 523
Ion Ratio Lower Upper
101 100
85 47.8 34.5 51.7
151 66.4 61.7 92.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#10
Methyl Iodide
Concen: 2.343 ug/l
RT: 4.082 min Scan# 371
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Tgt Ion:142 Resp: 4885
Ion Ratio Lower Upper
142 100
127 42.7 32.8 49.2
141 13.4 11.4 17.2





#11

Tert butyl alcohol

Concen: 19.495 ug/l

RT: 4.942 min Scan# 513

Delta R.T. -0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :





MDL-SOIL-03-QT2-2024

Tgt Ion: 59 Resp: 2664
 Ion Ratio Lower Upper
 59 100
 57 8.1 3.0 4.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
 Supervised By :Semsettin Yesilyurt 05/16/2024



#12

1,1-Dichloroethene

Concen: 3.085 ug/l




RT: 3.875 min Scan# 337


Delta R.T. 0.012 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20


Tgt Ion: 96 Resp: 5544
 Ion Ratio Lower Upper
 96 100
 61 109.7 114.6 171.8#
 98 48.0 50.8 76.2#





#13
Acrolein
Concen: 18.017 ug/l m
RT: 3.722 min Scan# 311
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024



Tgt Ion: 56 Resp: 265
Ion Ratio Lower Upper
56 100
55 80.8 55.7 83.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#14
Allyl chloride
Concen: 3.091 ug/l
RT: 4.472 min Scan# 435
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20


Tgt Ion: 41 Resp: 7066
Ion Ratio Lower Upper
41 100
39 77.0 58.3 87.5
76 45.7 39.0 58.6





#15
Acrylonitrile
Concen: 14.717 ug/l
RT: 5.161 min Scan# 546
Delta R.T. 0.012 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024



Tgt Ion: 53 Resp: 5219
Ion Ratio Lower Upper
53 100
52 87.8 64.7 97.1
51 40.1 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#16
Acetone
Concen: 17.205 ug/l
RT: 3.936 min Scan# 347
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20


Tgt Ion: 43 Resp: 4694
Ion Ratio Lower Upper
43 100
58 34.0 31.7 47.5





#17
Carbon Disulfide
Concen: 3.107 ug/l
RT: 4.186 min Scan# 388
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument : MSVOA_Y
ClientSampleId : MDL-SOIL-03-QT2-2024




Tgt Ion: 76 Resp: 1520
Ion Ratio Lower Upper
76 100
78 10.6 6.8 10.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#18
Methyl Acetate
Concen: 3.551 ug/l
RT: 4.460 min Scan# 433
Delta R.T. -0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Tgt Ion: 43 Resp: 2760
Ion Ratio Lower Upper
43 100
74 26.7 23.3 34.9





#19
Methyl tert-butyl Ether
Concen: 2.768 ug/l
RT: 5.216 min Scan# 511
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024




Tgt Ion: 73 Resp: 1108
Ion Ratio Lower Upper
73 100
57 23.2 17.3 25.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#20
Methylene Chloride
Concen: 6.146 ug/l
RT: 4.704 min Scan# 473
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Tgt Ion: 84 Resp: 11851
Ion Ratio Lower Upper
84 100
49 112.2 84.4 126.6
51 36.3 26.0 39.0
86 66.0 49.5 74.3





#21
trans-1,2-Dichloroethene
Concen: 3.112 ug/l
RT: 5.216 min Scan# 511
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024



Tgt Ion: 96 Resp: 5929
Ion Ratio Lower Upper
96 100
61 111.3 97.5 146.3
98 57.8 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#22
Diisopropyl ether
Concen: 2.967 ug/l
RT: 6.106 min Scan# 703
Delta R.T. -0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Tgt Ion: 45 Resp: 14389
Ion Ratio Lower Upper
45 100
43 55.3 42.2 63.2
87 40.9 25.0 37.6#
59 11.6 10.5 15.7









#23
 Vinyl Acetate
 Concen: 13.339 ug/l
 RT: 6.057 min Scan# 694
 Delta R.T. 0.006 min
 Lab File: VY018267.D
 Acq: 15 May 2024 11:20

Instrument :
 MSVOA_Y
 ClientSampleId :
 MDL-SOIL-03-QT2-2024

Tgt Ion: 43 Resp: 39319
 Ion Ratio Lower Upper
 43 100
 86 14.5 10.6 16.0




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
 Supervised By :Semsettin Yesilyurt 05/16/2024



#24
 1,1-Dichloroethane
 Concen: 2.991 ug/l
 RT: 6.003 min Scan# 686
 Delta R.T. -0.006 min
 Lab File: VY018267.D
 Acq: 15 May 2024 11:20


Tgt Ion: 63 Resp: 8936
 Ion Ratio Lower Upper
 63 100
 98 8.5 4.0 12.2
 100 3.7 2.6 8.0





#25
2-Butanone
Concen: 13.625 ug/l
RT: 6.978 min Scan# 846
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024




Tgt Ion: 43 Resp: 594
Ion Ratio Lower Upper
43 100
72 35.8 27.3 40.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#26
2,2-Dichloropropane
Concen: 2.990 ug/l
RT: 6.972 min Scan# 845
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Tgt Ion: 77 Resp: 8083
Ion Ratio Lower Upper
77 100
97 16.6 12.0 36.1





#27
 cis-1,2-Dichloroethene
 Concen: 3.045 ug/l
 RT: 6.984 min Scan# 846
 Delta R.T. 0.006 min
 Lab File: VY018267.D
 Acq: 15 May 2024 11:20




Instrument :
 MSVOA_Y
 ClientSampleId :
 MDL-SOIL-03-QT2-2024



Tgt Ion: 96 Resp: 685
 Ion Ratio Lower Upper
 96 100
 61 126.1 0.0 256.2
 98 66.2 0.0 130.0




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
 Supervised By :Semsettin Yesilyurt 05/16/2024



#28
 Bromochloromethane
 Concen: 3.221 ug/l
 RT: 7.332 min Scan# 904
 Delta R.T. -0.000 min
 Lab File: VY018267.D
 Acq: 15 May 2024 11:20


Tgt Ion: 49 Resp: 3846
 Ion Ratio Lower Upper
 49 100
 129 0.0 0.0 4.6
 130 85.5 72.7 109.1





#29
Tetrahydrofuran
Concen: 14.350 ug/l
RT: 7.350 min Scan# 907
Delta R.T. 0.012 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024



Tgt Ion: 42 Resp: 413
Ion Ratio Lower Upper
42 100
72 46.7 43.0 64.4
71 49.6 40.8 61.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#30
Chloroform
Concen: 3.025 ug/l
RT: 7.496 min Scan# 931
Delta R.T. -0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20


Tgt Ion: 83 Resp: 9433
Ion Ratio Lower Upper
83 100
85 62.6 52.2 78.4





#31
Cyclohexane
Concen: 4.080 ug/l
RT: 7.783 min Scan# 911
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024



Tgt Ion: 56 Resp: 1157
Ion Ratio Lower Upper
56 100
69 74.6 28.7 43.1
84 122.3 77.7 116.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#32
1,1,1-Trichloroethane
Concen: 2.826 ug/l
RT: 7.703 min Scan# 965
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Tgt Ion: 97 Resp: 7900
Ion Ratio Lower Upper
97 100
99 63.9 50.3 75.5
61 44.0 32.6 48.8









#33
1,2-Dichloroethane-d4
Concen: 29.885 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024

Tgt Ion: 65 Resp: 40850
Ion Ratio Lower Upper
65 100
67 55.5 0.0 109.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.691 min Scan# 1127
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20


Tgt Ion:114 Resp: 337173
Ion Ratio Lower Upper
114 100
63 16.3 0.0 35.0
88 13.8 0.0 27.2





#35
Dibromofluoromethane
Concen: 26.528 ug/l
RT: 7.716 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024




Tgt Ion: 113 Resp: 4700
Ion Ratio Lower Upper
113 100
111 101.0 82.2 123.4
192 19.6 15.9 23.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#36
1,1-Dichloropropene
Concen: 2.984 ug/l
RT: 7.917 min Scan# 1000
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Tgt Ion: 75 Resp: 7579
Ion Ratio Lower Upper
75 100
110 36.4 19.0 57.0
77 28.6 25.3 37.9





#37
Ethyl Acetate
Concen: 2.849 ug/l
RT: 7.069 min Scan# 80
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024



Tgt Ion: 43 Resp: 3040
Ion Ratio Lower Upper
43 100
61 15.5 12.6 19.0
70 14.5 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#38
Carbon Tetrachloride
Concen: 2.542 ug/l
RT: 7.899 min Scan# 997
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Tgt Ion:117 Resp: 7055
Ion Ratio Lower Upper
117 100
119 88.5 75.8 113.6
121 37.0 24.0 36.0





#39

Methylcyclohexane

Concen: 2.528 ug/l

RT: 9.185 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :





MDL-SOIL-03-QT2-2024

Tgt Ion: 83	Resp: 9264
Ion Ratio	Lower Upper
83 100	
55 64.2	51.5 77.3
98 54.7	40.5 60.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#40

Benzene

Concen: 2.693 ug/l




RT: 8.161 min Scan# 1040


Delta R.T. 0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20


Tgt Ion: 78	Resp: 22003
Ion Ratio	Lower Upper
78 100	
77 28.4	19.5 29.3





#41
Methacrylonitrile
Concen: 3.563 ug/l
RT: 7.313 min Scan# 901
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument : MSVOA_Y
Client Sample Id : MDL-SOIL-03-QT2-2024




Tgt Ion: 41 Resp: 2200
Ion Ratio Lower Upper
41 100
39 55.7 48.0 72.0
67 67.2 75.1 112.7
52 27.3 29.8 44.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#42
1,2-Dichloroethane
Concen: 2.892 ug/l
RT: 8.240 min Scan# 1053
Delta R.T. 0.012 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Tgt Ion: 62 Resp: 5141
Ion Ratio Lower Upper
62 100
98 5.7 0.0 22.0





#43
Isopropyl Acetate
Concen: 2.792 ug/l
RT: 8.270 min Scan# 1108
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument : MSVOA_Y
ClientSampleId : MDL-SOIL-03-QT2-2024



Tgt Ion: 43 Resp: 6118
Ion Ratio Lower Upper
43 100
61 35.3 27.9 41.9
87 17.5 13.1 19.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#44
Trichloroethene
Concen: 2.658 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Tgt Ion:130 Resp: 5913
Ion Ratio Lower Upper
130 100
95 100.8 0.0 188.0





#45

1,2-Dichloropropane

Concen: 2.749 ug/l

RT: 9.215 min Scan# 1213

Delta R.T. -0.000 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :

MDL-SOIL-03-QT2-2024

Tgt Ion: 63 Resp: 491

Ion Ratio Lower Upper

63 100





65 26.8 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024

Supervised By :Semsettin Yesilyurt 05/16/2024



#46

Dibromomethane

Concen: 2.593 ug/l

RT: 9.301 min Scan# 1227

Delta R.T. -0.000 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20




Tgt Ion: 93 Resp: 2663


Ion Ratio Lower Upper

93 100

95 82.8 66.2 99.4

174 102.9 80.9 121.3









#47
Bromodichloromethane
Concen: 2.493 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Instrument : MSVOA_Y
ClientSampleId : MDL-SOIL-03-QT2-2024

Tgt Ion: 83 Resp: 6330
Ion Ratio Lower Upper
83 100
85 60.7 50.1 75.1
127 7.3 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#48
Methyl methacrylate
Concen: 2.394 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20


Tgt Ion: 41 Resp: 2405
Ion Ratio Lower Upper
41 100
69 104.2 84.1 126.1
39 65.6 43.6 65.4





#49
1,4-Dioxane
Concen: 70.423 ug/l
RT: 9.289 min Scan# 1225
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024



Tgt Ion: 88 Resp: 960
Ion Ratio Lower Upper
88 100
43 30.5 0.0 0.0
58 58.6 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#50
Toluene-d8
Concen: 27.385 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20


Tgt Ion: 98 Resp: 191225
Ion Ratio Lower Upper
98 100
100 65.2 52.0 78.0





#51
4-Methyl-2-Pentanone
Concen: 12.176 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument : MSVOA_Y
ClientSampleId : MDL-SOIL-03-QT2-2024




Tgt Ion: 43 Resp: 1341
Ion Ratio Lower Upper
43 100
58 42.7 35.7 53.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#52
Toluene
Concen: 2.664 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Tgt Ion: 92 Resp: 14015
Ion Ratio Lower Upper
92 100
91 174.3 139.0 208.6





#53
t-1,3-Dichloropropene
Concen: 2.463 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument : MSVOA_Y
ClientSampleId : MDL-SOIL-03-QT2-2024



Tgt Ion: 75 Resp: 5840
Ion Ratio Lower Upper
75 100
77 30.9 24.1 36.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#54
cis-1,3-Dichloropropene
Concen: 2.557 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Tgt Ion: 75 Resp: 7413
Ion Ratio Lower Upper
75 100
77 28.2 25.4 38.0
39 34.8 29.8 44.6





#55

1,1,2-Trichloroethane

Concen: 2.448 ug/l

RT: 10.654 min Scan# 1449

Delta R.T. 0.012 min

Lab File: VY018267.D


Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :

MDL-SOIL-03-QT2-2024



Tgt Ion: 97 Resp: 345

Ion Ratio Lower Upper

97 100

83 58.7 67.5 101.3

85 36.8 42.9 64.3




99 65.1 46.6 70.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024

Supervised By :Semsettin Yesilyurt 05/16/2024



#56

Ethyl methacrylate

Concen: 2.196 ug/l

RT: 10.514 min Scan# 1426

Delta R.T. 0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20




Tgt Ion: 69 Resp: 4207


Ion Ratio Lower Upper

69 100

41 74.2 42.3 63.5

39 39.1 22.9 34.3





#57

1,3-Dichloropropane

Concen: 2.614 ug/l

RT: 10.794 min Scan# 1471

Delta R.T. 0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :

MDL-SOIL-03-QT2-2024

Tgt Ion: 76 Resp: 600

Ion Ratio Lower Upper

76 100





78 37.0 26.2 39.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024

Supervised By :Semsettin Yesilyurt 05/16/2024



#58

2-Chloroethyl Vinyl ether

Concen: 11.419 ug/l

RT: 9.782 min Scan# 1306

Delta R.T. -0.000 min

Lab File: VY018267.D




Acq: 15 May 2024 11:20


Tgt Ion: 63 Resp: 10447

Ion Ratio Lower Upper

63 100


106 32.5 23.7 35.5





#59
2-Hexanone
Concen: 10.648 ug/l
RT: 10.837 min Scan# 1478
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024



Tgt Ion: 43 Resp: 820
Ion Ratio Lower Upper
43 100
58 57.8 31.4 94.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#60
Dibromochloromethane
Concen: 2.281 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20


Tgt Ion:129 Resp: 4175
Ion Ratio Lower Upper
129 100
127 82.0 39.0 117.0





#61
1,2-Dibromoethane
Concen: 2.485 ug/l
RT: 11.093 min Scan# 1521
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024




Tgt Ion:107 Resp: 3302
Ion Ratio Lower Upper
107 100
109 88.7 75.9 113.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#62
4-Bromofluorobenzene
Concen: 24.954 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20



Tgt Ion: 95 Resp: 56271
Ion Ratio Lower Upper
95 100
174 85.0 0.0 175.6
176 82.1 0.0 171.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :

MDL-SOIL-03-QT2-2024





Tgt Ion:117 Resp: 27496

Ion	Ratio	Lower	Upper
117	100		
82	51.0	42.4	63.6
119	32.0	25.9	38.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#64

Tetrachloroethene

Concen: 2.592 ug/l

RT: 10.721 min Scan# 1460




Delta R.T. -0.000 min


Lab File: VY018267.D

Acq: 15 May 2024 11:20

Tgt Ion:164 Resp: 4745


Ion	Ratio	Lower	Upper
164	100		
166	114.1	102.6	154.0
129	99.0	72.0	108.0
131	90.0	70.0	105.0





#65
Chlorobenzene
Concen: 2.567 ug/l
RT: 11.514 min Scan# 1150
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument : MSVOA_Y
ClientSampleId : MDL-SOIL-03-QT2-2024




Tgt Ion:112 Resp: 1437
Ion Ratio Lower Upper
112 100
114 30.9 26.4 39.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 2.552 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20



Tgt Ion:131 Resp: 4622
Ion Ratio Lower Upper
131 100
133 88.7 47.8 143.3
119 77.2 34.1 102.2





#67

Ethyl Benzene

Concen: 2.690 ug/l

RT: 11.599 min Scan# 1604

Delta R.T. 0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :

MDL-SOIL-03-QT2-2024

Tgt Ion: 91 Resp: 25434

Ion Ratio Lower Upper

91 100





106 28.9 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024

Supervised By :Semsettin Yesilyurt 05/16/2024



#68

m/p-Xylenes

Concen: 5.320 ug/l

RT: 11.709 min Scan# 1622

Delta R.T. 0.006 min

Lab File: VY018267.D




Acq: 15 May 2024 11:20


Tgt Ion:106 Resp: 19892

Ion Ratio Lower Upper

106 100

91 192.6 156.0 234.0





#69

o-Xylene

Concen: 2.386 ug/l

RT: 12.032 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :

MDL-SOIL-03-QT2-2024

Tgt Ion:106 Resp: 856

Ion Ratio Lower Upper

106 100





91 223.9 103.6 310.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024

Supervised By :Semsettin Yesilyurt 05/16/2024



#70

Styrene

Concen: 2.498 ug/l

RT: 12.044 min Scan# 1677

Delta R.T. -0.000 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20




Tgt Ion:104 Resp: 14971


Ion Ratio Lower Upper

104 100

78 46.2 37.0 55.4

103 54.9 44.2 66.2





#71

Bromoform

Concen: 2.504 ug/l

RT: 12.209 min Scan# 1

Delta R.T. -0.000 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :





MDL-SOIL-03-QT2-2024

Tgt Ion:	173	Resp:	2510
Ion Ratio	Lower	Upper	
173	100		
175	47.0	25.1	75.3
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.428 min Scan# 1904


Delta R.T. 0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20


Tgt Ion:	152	Resp:	114541
Ion Ratio	Lower	Upper	
152	100		
115	57.2	28.2	84.7
150	158.0	0.0	345.6





#73
Isopropylbenzene
Concen: 2.744 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument : MSVOA_Y
ClientSampleId : MDL-SOIL-03-QT2-2024




Tgt Ion:105 Resp: 2356
Ion Ratio Lower Upper
105 100
120 29.3 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#74
N-ethyl acetate
Concen: 2.516 ug/l
RT: 12.154 min Scan# 1695
Delta R.T. 0.012 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20



Tgt Ion: 43 Resp: 4485
Ion Ratio Lower Upper
43 100
70 47.2 41.4 62.2
55 30.5 23.3 34.9
61 28.1 22.6 33.8





#75

1,1,2,2-Tetrachloroethane

Concen: 2.569 ug/l

RT: 12.581 min Scan# 1765

Delta R.T. -0.000 min

Lab File: VY018267.D


Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :

MDL-SOIL-03-QT2-2024






Tgt Ion: 83	Resp: 3520
Ion Ratio	Lower Upper
83	100
131	10.8 5.0 14.9
85	70.7 32.0 96.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#76

1,2,3-Trichloropropane

Concen: 2.403 ug/l m




RT: 12.642 min Scan# 1775


Delta R.T. 0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Tgt Ion: 75	Resp: 2203
Ion Ratio	Lower Upper
75	100
77	0.0 0.0 0.0









#77
Bromobenzene
Concen: 2.660 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024

Tgt Ion	Ratio	Lower	Upper
156	100		
77	175.5	87.1	261.3
158	96.6	48.6	145.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#78
n-propylbenzene
Concen: 2.868 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Tgt Ion	Ratio	Lower	Upper
91	100		
120	22.9	11.5	34.4









#79
2-Chlorotoluene
Concen: 2.819 ug/l
RT: 12.757 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024

Tgt Ion: 91 Resp: 15640
Ion Ratio Lower Upper
91 100
126 37.5 17.8 53.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#80
1,3,5-Trimethylbenzene
Concen: 2.690 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20


Tgt Ion:105 Resp: 18217
Ion Ratio Lower Upper
105 100
120 52.0 25.1 75.2





#81
trans-1,4-Dichloro-2-butene
Concen: 2.803 ug/l
RT: 12.379 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument : MSVOA_Y
ClientSampleId : MDL-SOIL-03-QT2-2024



Tgt Ion: 75 Resp: 1334
Ion Ratio Lower Upper
75 100
53 83.8 66.8 100.2
89 53.4 35.8 53.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#82
4-Chlorotoluene
Concen: 2.796 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Tgt Ion: 91 Resp: 15591
Ion Ratio Lower Upper
91 100
126 37.8 17.6 52.9





#83

tert-Butylbenzene

Concen: 2.807 ug/l

RT: 13.074 min Scan# 1846

Delta R.T. -0.000 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :





MDL-SOIL-03-QT2-2024

Tgt Ion:	119	Resp:	17530
Ion Ratio	Lower	Upper	
119	100		
91	60.5	31.3	93.8
134	27.7	13.4	40.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#84

1,2,4-Trimethylbenzene

Concen: 2.773 ug/l




RT: 13.123 min Scan# 1854


Delta R.T. 0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Tgt Ion:	105	Resp:	18396
Ion Ratio	Lower	Upper	
105	100		
120	46.0	23.2	69.6





#85

sec-Butylbenzene

Concen: 2.677 ug/l

RT: 13.257 min Scan# 1875

Delta R.T. 0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :





MDL-SOIL-03-QT2-2024

Tgt Ion:105 Resp: 2452
 Ion Ratio Lower Upper
 105 100
 134 19.3 10.1 30.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
 Supervised By :Semsettin Yesilyurt 05/16/2024



#86

p-Isopropyltoluene

Concen: 2.662 ug/l




RT: 13.373 min Scan# 1895


Delta R.T. 0.006 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20


Tgt Ion:119 Resp: 20061
 Ion Ratio Lower Upper
 119 100
 134 27.1 13.1 39.3
 91 21.9 10.7 32.1





#87
1,3-Dichlorobenzene
Concen: 2.624 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument : MSVOA_Y
ClientSampleId : MDL-SOIL-03-QT2-2024



Tgt Ion:146 Resp: 999
Ion Ratio Lower Upper
146 100
111 37.2 19.0 57.0
148 63.2 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#88
1,4-Dichlorobenzene
Concen: 2.701 ug/l m
RT: 13.446 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Tgt Ion:146 Resp: 10032
Ion Ratio Lower Upper
146 100
111 37.0 18.4 55.4
148 63.0 32.4 97.2









#89
n-Butylbenzene
Concen: 2.611 ug/l
RT: 13.702 min Scan# 1949
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20

Instrument :
MSVOA_Y
ClientSampleId :
MDL-SOIL-03-QT2-2024

Tgt Ion: 91 Resp: 17809
Ion Ratio Lower Upper
91 100
92 53.8 26.6 79.7
134 26.0 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#90
Hexachloroethane
Concen: 2.485 ug/l
RT: 13.958 min Scan# 1991
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20


Tgt Ion:117 Resp: 3674
Ion Ratio Lower Upper
117 100
201 67.3 37.4 112.2





#91
1,2-Dichlorobenzene
Concen: 2.583 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. -0.000 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20




Instrument : MSVOA_Y
ClientSampleId : MDL-SOIL-03-QT2-2024




Tgt Ion:146 Resp: 846
Ion Ratio Lower Upper
146 100
111 42.8 19.9 59.7
148 59.6 31.8 95.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 2.230 ug/l
RT: 14.361 min Scan# 2057
Delta R.T. 0.006 min
Lab File: VY018267.D
Acq: 15 May 2024 11:20



Tgt Ion: 75 Resp: 416
Ion Ratio Lower Upper
75 100
155 73.1 48.6 145.8
157 114.7 62.7 188.2





#93

1,2,4-Trichlorobenzene

Concen: 2.344 ug/l

RT: 15.007 min Scan# 2163

Delta R.T. -0.000 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :





MDL-SOIL-03-QT2-2024

Tgt Ion:180	Resp:	431
Ion Ratio	Lower	Upper
180	100	
182	88.3	47.9 143.8
145	31.2	14.2 42.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#94

Hexachlorobutadiene

Concen: 2.314 ug/l




RT: 15.123 min Scan# 2182


Delta R.T. 0.012 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Tgt Ion:225	Resp:	2351
Ion Ratio	Lower	Upper
225	100	
223	68.3	30.4 91.2
227	63.7	31.9 95.9





#95

Naphthalene

Concen: 2.111 ug/l

RT: 15.239 min Scan# 2200

Delta R.T. 0.006 min

Lab File: VY018267.D


Acq: 15 May 2024 11:20

Instrument :

MSVOA_Y

ClientSampleId :

MDL-SOIL-03-QT2-2024






Tgt Ion	Ratio	Lower	Upper
128	100		
127	13.9	10.6	16.0
129	10.0	8.6	13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2024
Supervised By :Semsettin Yesilyurt 05/16/2024



#96

1,2,3-Trichlorobenzene

Concen: 2.256 ug/l

RT: 15.434 min Scan# 2233

Delta R.T. 0.012 min

Lab File: VY018267.D

Acq: 15 May 2024 11:20

Tgt Ion	Ratio	Lower	Upper
180	100		
182	92.9	47.6	142.9
145	33.7	15.4	46.1

