



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051524\
Data File : VY018264.D
Acq On : 15 May 2024 09:03
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Title : SW846 8260
Last Update : Wed May 08 01:40:14 2024



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.3	23571	PASS
75	95	30	60	49.7	64069	PASS
95	95	100	100	100.0	128808	PASS
96	95	5	9	6.4	8218	PASS
173	174	0.00	2	1.0	995	PASS
174	95	50	100	81.0	104365	PASS
175	174	5	9	7.2	7495	PASS
176	174	95	101	98.1	102379	PASS
177	176	5	9	6.6	6791	PASS