

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
 Data File : VY022256.D
 Acq On : 15 May 2025 11:02
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
 Supervised By :Semsettin Yesilyurt 05/16/2025

Quant Time: May 16 01:23:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:19:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	267598	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	439323	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	377571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	180358	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	144713	49.481	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	98.960%		
35) Dibromofluoromethane	7.634	113	132404	50.495	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	100.980%		
50) Toluene-d8	10.103	98	540296	50.311	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	100.620%		
62) 4-Bromofluorobenzene	12.408	95	169827	49.891	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	99.780%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.861	85	135726	48.608	ug/l	100
3) Chloromethane	2.062	50	278892	47.481	ug/l	100
4) Vinyl Chloride	2.202	62	323792	49.341	ug/l	100
5) Bromomethane	2.592	94	234736	43.941	ug/l	100
6) Chloroethane	2.732	64	202370	49.569	ug/l	100
7) Trichlorofluoromethane	3.050	101	324887	48.898	ug/l	100
8) Diethyl Ether	3.452	74	78670	49.338	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.818	101	144018	49.930	ug/l	100
10) Methyl Iodide	4.007	142	183918	55.091	ug/l	100
11) Tert butyl alcohol	4.872	59	51598	231.036	ug/l #	100
12) 1,1-Dichloroethene	3.787	96	144103	50.344	ug/l	100
13) Acrolein	3.653	56	75423	231.010	ug/l	100
14) Allyl chloride	4.379	41	256499	51.106	ug/l	100
15) Acrylonitrile	5.061	53	160496	244.283	ug/l	100
16) Acetone	3.873	43	124421	226.888	ug/l	100
17) Carbon Disulfide	4.104	76	468979	50.244	ug/l	100
18) Methyl Acetate	4.385	43	80820	48.436	ug/l	100
19) Methyl tert-butyl Ether	5.116	73	390412	49.466	ug/l	100
20) Methylene Chloride	4.616	84	159764	45.087	ug/l	100
21) trans-1,2-Dichloroethene	5.110	96	158544	50.181	ug/l	100
22) Diisopropyl ether	6.019	45	531560	50.152	ug/l	100
23) Vinyl Acetate	5.958	43	1570214	245.223	ug/l	100
24) 1,1-Dichloroethane	5.915	63	298823	50.465	ug/l	100
25) 2-Butanone	6.896	43	215828	235.291	ug/l	100
26) 2,2-Dichloropropane	6.884	77	264141	50.153	ug/l	100
27) cis-1,2-Dichloroethene	6.890	96	184842	50.698	ug/l	100
28) Bromochloromethane	7.244	49	126541	50.100	ug/l	100
29) Tetrahydrofuran	7.262	42	142772	241.931	ug/l	100
30) Chloroform	7.421	83	288708	50.831	ug/l	100
31) Cyclohexane	7.701	56	282375	47.728	ug/l	100
32) 1,1,1-Trichloroethane	7.616	97	259382	49.777	ug/l	100
36) 1,1-Dichloropropene	7.835	75	221943	51.388	ug/l	100
37) Ethyl Acetate	6.988	43	102924	49.529	ug/l	100
38) Carbon Tetrachloride	7.817	117	226903	50.754	ug/l	100
39) Methylcyclohexane	9.109	83	294336	50.726	ug/l	100
40) Benzene	8.079	78	658644	51.442	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
 Data File : VY022256.D
 Acq On : 15 May 2025 11:02
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
 Supervised By :Semsettin Yesilyurt 05/16/2025

Quant Time: May 16 01:23:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:19:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.220	41	69476m	45.691	ug/l	
42) 1,2-Dichloroethane	8.158	62	176425	50.954	ug/l	100
43) Isopropyl Acetate	8.195	43	225628	49.785	ug/l	100
44) Trichloroethene	8.866	130	162802	51.138	ug/l	100
45) 1,2-Dichloropropane	9.140	63	156870	50.819	ug/l	100
46) Dibromomethane	9.231	93	82994	49.849	ug/l	100
47) Bromodichloromethane	9.420	83	222159	51.435	ug/l	100
48) Methyl methacrylate	9.219	41	105390	49.726	ug/l	100
49) 1,4-Dioxane	9.231	88	18027	958.446	ug/l	100
51) 4-Methyl-2-Pentanone	9.999	43	556123	250.085	ug/l	100
52) Toluene	10.170	92	414584	51.134	ug/l	100
53) t-1,3-Dichloropropene	10.390	75	213993	50.768	ug/l	100
54) cis-1,3-Dichloropropene	9.853	75	246990	51.143	ug/l	100
55) 1,1,2-Trichloroethane	10.573	97	106676	50.036	ug/l	100
56) Ethyl methacrylate	10.438	69	167259	50.364	ug/l	100
57) 1,3-Dichloropropane	10.719	76	192275	50.759	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	403074	270.408	ug/l	100
59) 2-Hexanone	10.762	43	365111	244.881	ug/l	100
60) Dibromochloromethane	10.908	129	141896	51.278	ug/l	100
61) 1,2-Dibromoethane	11.012	107	99822	50.236	ug/l	100
64) Tetrachloroethene	10.646	164	178854	51.249	ug/l	100
65) Chlorobenzene	11.438	112	435874	51.169	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.518	131	148798	50.811	ug/l	100
67) Ethyl Benzene	11.518	91	821168	51.473	ug/l	100
68) m/p-Xylenes	11.627	106	617646	102.746	ug/l	100
69) o-Xylene	11.950	106	289897	50.971	ug/l	100
70) Styrene	11.969	104	491792	51.260	ug/l	100
71) Bromoform	12.133	173	78794	49.849	ug/l #	100
73) Isopropylbenzene	12.255	105	756905	50.657	ug/l	100
74) N-amyl acetate	12.066	43	209097	48.930	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.505	83	113234	49.277	ug/l	100
76) 1,2,3-Trichloropropane	12.554	75	90281m	28.402	ug/l	
77) Bromobenzene	12.530	156	161475	50.599	ug/l	100
78) n-propylbenzene	12.597	91	924664	51.306	ug/l	100
79) 2-Chlorotoluene	12.676	91	505684	51.058	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	615033	50.991	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	42692	47.331	ug/l	100
82) 4-Chlorotoluene	12.773	91	524270	51.464	ug/l	100
83) tert-Butylbenzene	12.999	119	546231	50.952	ug/l	100
84) 1,2,4-Trimethylbenzene	13.042	105	615550	51.410	ug/l	100
85) sec-Butylbenzene	13.176	105	824725	51.825	ug/l	100
86) p-Isopropyltoluene	13.292	119	688561	51.446	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	331536	50.890	ug/l	100
88) 1,4-Dichlorobenzene	13.365	146	315598	50.363	ug/l	100
89) n-Butylbenzene	13.615	91	661532	52.245	ug/l	100
90) Hexachloroethane	13.877	117	137225	50.791	ug/l	100
91) 1,2-Dichlorobenzene	13.657	146	276666	50.558	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	18039	48.682	ug/l	100
93) 1,2,4-Trichlorobenzene	14.919	180	162275	51.338	ug/l	100
94) Hexachlorobutadiene	15.023	225	90452	51.313	ug/l	100
95) Naphthalene	15.145	128	298075	49.692	ug/l	100
96) 1,2,3-Trichlorobenzene	15.328	180	135673	50.596	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
Data File : VY022256.D
Acq On : 15 May 2025 11:02
Operator : SY/MD
Sample : VSTDICCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025

Quant Time: May 16 01:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:19:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
Data File : VY022256.D
Acq On : 15 May 2025 11:02
Operator : SY/MD
Sample : VSTDICCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050

Quant Time: May 16 01:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:19:39 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025

