

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
 Data File : VY022258.D
 Acq On : 15 May 2025 11:47
 Operator : SY/MD
 Sample : VSTDICC150
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
 Supervised By :Semsettin Yesilyurt 05/16/2025

Quant Time: May 16 01:25:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:19:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	271027	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	442917	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	389769	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	187978	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	455385	153.739	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 307.480%#			
35) Dibromofluoromethane	7.634	113	412318	155.969	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 311.940%#			
50) Toluene-d8	10.103	98	1689714	156.065	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 312.120%#			
62) 4-Bromofluorobenzene	12.402	95	541367	157.751	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 315.500%#			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.861	85	389593	137.760	ug/l	100
3) Chloromethane	2.068	50	727243	122.246	ug/l	99
4) Vinyl Chloride	2.202	62	902596	135.801	ug/l	97
5) Bromomethane	2.586	94	780107	144.182	ug/l	99
6) Chloroethane	2.727	64	590870	142.898	ug/l	98
7) Trichlorofluoromethane	3.050	101	989291	147.013	ug/l	99
8) Diethyl Ether	3.452	74	236216	146.270	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.812	101	422397	144.588	ug/l	100
10) Methyl Iodide	4.001	142	518980	153.489	ug/l	98
11) Tert butyl alcohol	4.866	59	172549	762.834	ug/l #	100
12) 1,1-Dichloroethene	3.787	96	425285	146.697	ug/l	97
13) Acrolein	3.653	56	237291	717.595	ug/l	98
14) Allyl chloride	4.379	41	756728	148.866	ug/l	100
15) Acrylonitrile	5.055	53	512850	770.708	ug/l	100
16) Acetone	3.873	43	375443	675.976	ug/l	99
17) Carbon Disulfide	4.104	76	1365378	144.429	ug/l	100
18) Methyl Acetate	4.385	43	282084	166.915	ug/l	99
19) Methyl tert-butyl Ether	5.116	73	1225258	153.279	ug/l	98
20) Methylene Chloride	4.616	84	448817	125.059	ug/l	99
21) trans-1,2-Dichloroethene	5.110	96	472143	147.547	ug/l	99
22) Diisopropyl ether	6.019	45	1626927	151.557	ug/l	99
23) Vinyl Acetate	5.958	43	4953644	763.831	ug/l	100
24) 1,1-Dichloroethane	5.915	63	879785	146.697	ug/l	99
25) 2-Butanone	6.890	43	708010	762.091	ug/l	98
26) 2,2-Dichloropropane	6.884	77	787195	147.575	ug/l	100
27) cis-1,2-Dichloroethene	6.890	96	563503	152.601	ug/l	100
28) Bromochloromethane	7.244	49	387221	151.369	ug/l	100
29) Tetrahydrofuran	7.262	42	472224	790.072	ug/l	100
30) Chloroform	7.421	83	859317	149.381	ug/l	98
31) Cyclohexane	7.701	56	821784	137.144	ug/l	98
32) 1,1,1-Trichloroethane	7.616	97	784014	148.554	ug/l	99
36) 1,1-Dichloropropene	7.835	75	658871	151.316	ug/l	99
37) Ethyl Acetate	6.982	43	320686	153.067	ug/l	99
38) Carbon Tetrachloride	7.817	117	696139	154.450	ug/l	100
39) Methylcyclohexane	9.110	83	886329	151.510	ug/l	98
40) Benzene	8.079	78	1987535	153.973	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
 Data File : VY022258.D
 Acq On : 15 May 2025 11:47
 Operator : SY/MD
 Sample : VSTDICC150
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
 Supervised By :Semsettin Yesilyurt 05/16/2025

Quant Time: May 16 01:25:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:19:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.220	41	202603m	132.161	ug/l	
42) 1,2-Dichloroethane	8.159	62	534517	153.124	ug/l	99
43) Isopropyl Acetate	8.195	43	704366	154.159	ug/l #	89
44) Trichloroethene	8.866	130	481990	150.171	ug/l	96
45) 1,2-Dichloropropane	9.140	63	472217	151.736	ug/l	97
46) Dibromomethane	9.231	93	258087	153.757	ug/l	99
47) Bromodichloromethane	9.420	83	669834	153.824	ug/l	99
48) Methyl methacrylate	9.219	41	344147	161.060	ug/l	99
49) 1,4-Dioxane	9.231	88	61019	3217.888	ug/l	95
51) 4-Methyl-2-Pentanone	10.000	43	1829015	815.822	ug/l	99
52) Toluene	10.170	92	1276917	156.216	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	663513	156.134	ug/l	100
54) cis-1,3-Dichloropropene	9.853	75	754493	154.963	ug/l	99
55) 1,1,2-Trichloroethane	10.573	97	330716	153.863	ug/l	99
56) Ethyl methacrylate	10.439	69	535897	160.058	ug/l	99
57) 1,3-Dichloropropane	10.719	76	584253	152.987	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.713	63	1350536	898.675	ug/l	99
59) 2-Hexanone	10.762	43	1227881	816.860	ug/l	99
60) Dibromochloromethane	10.908	129	438666	157.237	ug/l	99
61) 1,2-Dibromoethane	11.012	107	308529	154.008	ug/l	99
64) Tetrachloroethene	10.646	164	501627	139.237	ug/l	100
65) Chlorobenzene	11.438	112	1336594	151.997	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.518	131	480525	158.952	ug/l	100
67) Ethyl Benzene	11.518	91	2598027	157.756	ug/l	99
68) m/p-Xylenes	11.627	106	1992413	321.066	ug/l	99
69) o-Xylene	11.957	106	937727	159.714	ug/l	99
70) Styrene	11.969	104	1613798	162.944	ug/l	100
71) Bromoform	12.133	173	254251	155.818	ug/l #	99
73) Isopropylbenzene	12.255	105	2357568	151.387	ug/l	100
74) N-amyl acetate	12.066	43	694999	156.041	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.505	83	375378	156.735	ug/l	100
76) 1,2,3-Trichloropropane	12.554	75	289968m	87.525	ug/l	
77) Bromobenzene	12.530	156	509579	153.206	ug/l	99
78) n-propylbenzene	12.591	91	2822794	150.276	ug/l	100
79) 2-Chlorotoluene	12.676	91	1563003	151.415	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	1888432	150.219	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.304	75	139931	148.846	ug/l	99
82) 4-Chlorotoluene	12.774	91	1616520	152.250	ug/l	99
83) tert-Butylbenzene	12.993	119	1699962	152.144	ug/l	100
84) 1,2,4-Trimethylbenzene	13.042	105	1934641	155.030	ug/l	100
85) sec-Butylbenzene	13.176	105	2493393	150.330	ug/l	100
86) p-Isopropyltoluene	13.292	119	2205275	158.088	ug/l	99
87) 1,3-Dichlorobenzene	13.286	146	1079479	158.981	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	969874	148.497	ug/l	99
89) n-Butylbenzene	13.615	91	1992554	150.985	ug/l	99
90) Hexachloroethane	13.877	117	417341	148.208	ug/l	99
91) 1,2-Dichlorobenzene	13.658	146	858733	150.565	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	58383	151.172	ug/l	99
93) 1,2,4-Trichlorobenzene	14.919	180	508537	154.361	ug/l	99
94) Hexachlorobutadiene	15.023	225	261832	142.514	ug/l	100
95) Naphthalene	15.139	128	979369	156.653	ug/l	100
96) 1,2,3-Trichlorobenzene	15.328	180	426231	152.509	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
Data File : VY022258.D
Acq On : 15 May 2025 11:47
Operator : SY/MD
Sample : VSTDICC150
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025

Quant Time: May 16 01:25:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:19:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
Data File : VY022258.D
Acq On : 15 May 2025 11:47
Operator : SY/MD
Sample : VSTDICC150
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025

Quant Time: May 16 01:25:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:19:39 2025
Response via : Initial Calibration

