

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
 Data File : VY022259.D
 Acq On : 15 May 2025 12:34
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: May 16 01:46:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:42:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	370240	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	643452	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	512184	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	186798	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	205860	50.875	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.760%
35) Dibromofluoromethane	7.634	113	193643	50.421	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.840%
50) Toluene-d8	10.109	98	769774	48.940	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.880%
62) 4-Bromofluorobenzene	12.408	95	205688	41.257	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
Data File : VY022259.D
Acq On : 15 May 2025 12:34
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: May 16 01:46:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:42:09 2025
Response via : Initial Calibration

