

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
 Data File : VY022265.D
 Acq On : 15 May 2025 15:35
 Operator : SY/MD
 Sample : Q1984-03
 Misc : 5.65g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-PCS-TC-34-050725

Quant Time: May 16 01:56:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:42:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	328781	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	578595	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	473987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	189383	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	170135	47.348	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.700%
35) Dibromofluoromethane	7.634	113	170320	49.320	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.640%
50) Toluene-d8	10.103	98	702946	49.701	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.400%
62) 4-Bromofluorobenzene	12.401	95	194312	43.344	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.680%
Target Compounds						
16) Acetone	3.866	43	15631	23.200	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
Data File : VY022265.D
Acq On : 15 May 2025 15:35
Operator : SY/MD
Sample : Q1984-03
Misc : 5.65g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-PCS-TC-34-050725

Quant Time: May 16 01:56:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:42:09 2025
Response via : Initial Calibration

