

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051624\
 Data File : VY018290.D
 Acq On : 16 May 2024 17:11
 Operator : SY/MD
 Sample : P2497-07
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 280646

Quant Time: May 17 01:17:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	137680	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	273697	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	245671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	90153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	76348	75.378	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 150.760%			
35) Dibromofluoromethane	7.710	113	82726	57.512	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 115.020%			
50) Toluene-d8	10.179	98	318854	56.253	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 112.500%			
62) 4-Bromofluorobenzene	12.483	95	95175	51.994	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 103.980%			
Target Compounds						
16) Acetone	3.936	43	3646	18.038	ug/l	96
20) Methylene Chloride	4.704	84	14071	9.849	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051624\
Data File : VY018290.D
Acq On : 16 May 2024 17:11
Operator : SY/MD
Sample : P2497-07
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
280646

Quant Time: May 17 01:17:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

