

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051624\
 Data File : VY018291.D
 Acq On : 16 May 2024 17:34
 Operator : SY/MD
 Sample : P2497-05
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT3997

Quant Time: May 17 01:17:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	136336	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	277276	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	244443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	86765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	77711	77.480	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	154.960%
35) Dibromofluoromethane	7.716	113	82516	56.626	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.260%
50) Toluene-d8	10.179	98	318777	55.513	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	111.020%
62) 4-Bromofluorobenzene	12.483	95	95224	51.349	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	102.700%
Target Compounds						
						Qvalue
16) Acetone	3.936	43	4765	23.807	ug/l	91
20) Methylene Chloride	4.704	84	12623	8.923	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051624\
Data File : VY018291.D
Acq On : 16 May 2024 17:34
Operator : SY/MD
Sample : P2497-05
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT3997

Quant Time: May 17 01:17:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

