

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051624\
 Data File : VY018293.D
 Acq On : 16 May 2024 18:22
 Operator : SY/MD
 Sample : P2513-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-13

Quant Time: May 17 01:18:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 01:40:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	136306	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	270899	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	238154	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	88782	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	70979	70.784	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.560%
35) Dibromofluoromethane	7.710	113	79989	56.184	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.360%
50) Toluene-d8	10.179	98	316291	56.377	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.760%
62) 4-Bromofluorobenzene	12.477	95	93110	51.391	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	102.780%
Target Compounds						
16) Acetone	3.948	43	2116	10.574	ug/l #	73
20) Methylene Chloride	4.704	84	16698	11.806	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051624\
Data File : VY018293.D
Acq On : 16 May 2024 18:22
Operator : SY/MD
Sample : P2513-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-13

Quant Time: May 17 01:18:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Quant Title : SW846 8260
QLast Update : Wed May 08 01:40:14 2024
Response via : Initial Calibration

