



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051624\
Data File : VY018281.D
Acq On : 16 May 2024 08:25
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050724S.M
Title : SW846 8260
Last Update : Wed May 08 01:40:14 2024



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.4	21929	PASS
75	95	30	60	50.5	60235	PASS
95	95	100	100	100.0	119261	PASS
96	95	5	9	7.0	8292	PASS
173	174	0.00	2	0.9	849	PASS
174	95	50	100	80.5	95971	PASS
175	174	5	9	7.1	6810	PASS
176	174	95	101	97.1	93147	PASS
177	176	5	9	6.5	6037	PASS