

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051625\
 Data File : VY022284.D
 Acq On : 16 May 2025 12:29
 Operator : SY/MD
 Sample : Q1984-13
 Misc : 3.42g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-27-050725

Quant Time: May 17 01:31:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:42:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	290740	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	521620	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	417665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	152557	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	156001	49.095	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.200%
35) Dibromofluoromethane	7.634	113	154975	49.778	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.560%
50) Toluene-d8	10.109	98	624035	48.941	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.880%
62) 4-Bromofluorobenzene	12.402	95	162278	40.152	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.300%
Target Compounds						Qvalue
20) Methylene Chloride	4.610	84	9836	2.555	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051625\
Data File : VY022284.D
Acq On : 16 May 2025 12:29
Operator : SY/MD
Sample : Q1984-13
Misc : 3.42g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-27-050725

Quant Time: May 17 01:31:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:42:09 2025
Response via : Initial Calibration

