

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051625\
 Data File : VY022301.D
 Acq On : 16 May 2025 19:29
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : Mahesh Dadoda 05/19/2025
 Supervised By : Semsettin Yesilyurt 05/19/2025

Quant Time: May 17 01:40:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:42:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	169408	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	301400	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	262982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	124169	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	105132	56.783	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.560%	
35) Dibromofluoromethane	7.634	113	95967	53.347	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.700%	
50) Toluene-d8	10.103	98	385048	52.262	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 104.520%	
62) 4-Bromofluorobenzene	12.401	95	118638	50.802	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 101.600%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.861	85	81891	46.326	ug/l	98
3) Chloromethane	2.068	50	194341	52.264	ug/l	100
4) Vinyl Chloride	2.202	62	261518	62.949	ug/l	97
5) Bromomethane	2.592	94	216391	63.984	ug/l	99
6) Chloroethane	2.732	64	182258	70.518	ug/l	98
7) Trichlorofluoromethane	3.049	101	229841	54.643	ug/l	99
8) Diethyl Ether	3.452	74	57553	57.015	ug/l	94
9) 1,1,2-Trichlorotrifluo...	3.818	101	91858	50.304	ug/l	98
10) Methyl Iodide	4.000	142	124272	58.800	ug/l	99
11) Tert butyl alcohol	4.866	59	41446	293.143	ug/l #	100
12) 1,1-Dichloroethene	3.787	96	96302	53.144	ug/l	99
13) Acrolein	3.647	56	45853	221.842	ug/l	97
14) Allyl chloride	4.378	41	160107	50.390	ug/l	97
15) Acrylonitrile	5.055	53	122379	294.229	ug/l	99
16) Acetone	3.872	43	90742	261.381	ug/l	98
17) Carbon Disulfide	4.104	76	310118	52.482	ug/l	98
18) Methyl Acetate	4.385	43	78917	72.098	ug/l	98
19) Methyl tert-butyl Ether	5.110	73	289508	57.942	ug/l	96
20) Methylene Chloride	4.610	84	112252	50.040	ug/l	97
21) trans-1,2-Dichloroethene	5.110	96	107400	53.696	ug/l	98
22) Diisopropyl ether	6.012	45	357394	53.264	ug/l	96
23) Vinyl Acetate	5.957	43	1023992	252.608	ug/l	98
24) 1,1-Dichloroethane	5.909	63	202421	53.998	ug/l	98
25) 2-Butanone	6.890	43	159016	273.834	ug/l	99
26) 2,2-Dichloropropane	6.878	77	154549	46.353	ug/l	98
27) cis-1,2-Dichloroethene	6.884	96	129025	55.900	ug/l	96
28) Bromochloromethane	7.244	49	87604	54.788	ug/l	97
29) Tetrahydrofuran	7.262	42	108553	290.562	ug/l	96
30) Chloroform	7.421	83	202440	56.301	ug/l	96
31) Cyclohexane	7.695	56	171539	45.799	ug/l	97
32) 1,1,1-Trichloroethane	7.616	97	172530	52.300	ug/l	100
36) 1,1-Dichloropropene	7.835	75	143691	48.494	ug/l	99
37) Ethyl Acetate	6.982	43	71726	50.310	ug/l	97
38) Carbon Tetrachloride	7.817	117	147865	48.210	ug/l	99
39) Methylcyclohexane	9.109	83	180031	45.224	ug/l	97
40) Benzene	8.079	78	456109	51.925	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051625\
 Data File : VY022301.D
 Acq On : 16 May 2025 19:29
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025
 Supervised By :Semsettin Yesilyurt 05/19/2025

Quant Time: May 17 01:40:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:42:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.213	41	39846	42.134	ug/l #	77
42) 1,2-Dichloroethane	8.152	62	126733	53.352	ug/l	99
43) Isopropyl Acetate	8.195	43	161047	51.797	ug/l #	89
44) Trichloroethene	8.865	130	110550	50.616	ug/l	95
45) 1,2-Dichloropropane	9.140	63	108718	51.337	ug/l	97
46) Dibromomethane	9.231	93	61598	53.928	ug/l	98
47) Bromodichloromethane	9.420	83	155944	52.627	ug/l	97
48) Methyl methacrylate	9.219	41	75248	51.751	ug/l	94
49) 1,4-Dioxane	9.237	88	14407	1116.500	ug/l #	86
51) 4-Methyl-2-Pentanone	9.993	43	412211	270.195	ug/l	98
52) Toluene	10.170	92	284515	51.150	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	147108	50.870	ug/l	99
54) cis-1,3-Dichloropropene	9.853	75	169256	51.085	ug/l	98
55) 1,1,2-Trichloroethane	10.572	97	78105	53.400	ug/l	98
56) Ethyl methacrylate	10.438	69	120471	52.876	ug/l	99
57) 1,3-Dichloropropane	10.713	76	138601	53.333	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.707	63	300672	294.015	ug/l	100
59) 2-Hexanone	10.755	43	270964	264.900	ug/l	98
60) Dibromochloromethane	10.908	129	101289	53.353	ug/l	99
61) 1,2-Dibromoethane	11.011	107	73769	54.113	ug/l	100
64) Tetrachloroethene	10.646	164	117020	48.141	ug/l	96
65) Chlorobenzene	11.438	112	298366	50.288	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.511	131	103057	50.525	ug/l	99
67) Ethyl Benzene	11.517	91	552037	49.681	ug/l	98
68) m/p-Xylenes	11.627	106	416782	99.542	ug/l	99
69) o-Xylene	11.950	106	200603	50.639	ug/l	99
70) Styrene	11.969	104	342456	51.248	ug/l	98
71) Bromoform	12.133	173	58041	52.719	ug/l #	99
73) Isopropylbenzene	12.249	105	506442	49.232	ug/l	100
74) N-amyl acetate	12.066	43	145082	49.313	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.505	83	86108	54.429	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	62208m	49.451	ug/l	
77) Bromobenzene	12.529	156	111837	50.903	ug/l	100
78) n-propylbenzene	12.590	91	615206	49.582	ug/l	99
79) 2-Chlorotoluene	12.676	91	339182	49.744	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	405753	48.863	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.298	75	28728	46.262	ug/l	100
82) 4-Chlorotoluene	12.773	91	345271	49.230	ug/l	98
83) tert-Butylbenzene	12.993	119	369198	50.023	ug/l	98
84) 1,2,4-Trimethylbenzene	13.042	105	411972	49.978	ug/l	98
85) sec-Butylbenzene	13.176	105	538476	49.149	ug/l	100
86) p-Isopropyltoluene	13.291	119	449798	48.814	ug/l	99
87) 1,3-Dichlorobenzene	13.285	146	224283	50.006	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	217882	50.503	ug/l	98
89) n-Butylbenzene	13.615	91	423157	48.542	ug/l	99
90) Hexachloroethane	13.877	117	91050	48.950	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	193609	51.391	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	13510	52.958	ug/l	99
93) 1,2,4-Trichlorobenzene	14.919	180	107697	49.489	ug/l	100
94) Hexachlorobutadiene	15.023	225	56596	46.635	ug/l	98
95) Naphthalene	15.139	128	217144	52.582	ug/l	100
96) 1,2,3-Trichlorobenzene	15.328	180	93303	50.540	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051625\
Data File : VY022301.D
Acq On : 16 May 2025 19:29
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025

Quant Time: May 17 01:40:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:42:09 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051625\
 Data File : VY022301.D
 Acq On : 16 May 2025 19:29
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Quant Time: May 17 01:40:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:42:09 2025
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
 Supervised By :Semsettin Yesilyurt 05/19/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 169408
Ion Ratio Lower Upper
168 100
99 54.6 44.2 66.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#2
Dichlorodifluoromethane
Concen: 46.326 ug/l
RT: 1.861 min Scan# 23
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29


Tgt Ion: 85 Resp: 81891
Ion Ratio Lower Upper
85 100
87 32.1 16.5 49.5





#3
Chloromethane
Concen: 52.264 ug/l
RT: 2.068 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 50 Resp: 19434
Ion Ratio Lower Upper
50 100
52 33.0 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#4
Vinyl Chloride
Concen: 62.949 ug/l
RT: 2.202 min Scan# 79
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Tgt Ion: 62 Resp: 261518
Ion Ratio Lower Upper
62 100
64 32.3 24.5 36.7





#5
Bromomethane
Concen: 63.984 ug/l
RT: 2.592 min Scan# 143
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 94 Resp: 21639
Ion Ratio Lower Upper
94 100
96 92.7 73.4 110.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#6
Chloroethane
Concen: 70.518 ug/l
RT: 2.732 min Scan# 166
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Tgt Ion: 64 Resp: 182258
Ion Ratio Lower Upper
64 100
66 31.9 24.8 37.2





#7
 Trichlorofluoromethane
 Concen: 54.643 ug/l
 RT: 3.049 min Scan# 218
 Delta R.T. -0.000 min
 Lab File: VY022301.D
 Acq: 16 May 2025 19:29




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion:101 Resp: 22984
 Ion Ratio Lower Upper
 101 100
 103 64.1 51.7 77.5




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
 Supervised By :Semsettin Yesilyurt 05/19/2025



#8
 Diethyl Ether
 Concen: 57.015 ug/l
 RT: 3.452 min Scan# 284
 Delta R.T. -0.000 min
 Lab File: VY022301.D
 Acq: 16 May 2025 19:29

Tgt Ion: 74 Resp: 57553
 Ion Ratio Lower Upper
 74 100
 45 91.3 48.6 145.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.304 ug/l
RT: 3.818 min Scan# 344
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 91858
Ion Ratio Lower Upper
101 100
85 48.3 38.6 58.0
151 72.4 59.7 89.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#10
Methyl Iodide
Concen: 58.800 ug/l
RT: 4.000 min Scan# 374
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29


Tgt Ion:142 Resp: 124272
Ion Ratio Lower Upper
142 100
127 44.9 35.7 53.5
141 14.6 11.1 16.7





#11
Tert butyl alcohol
Concen: 293.143 ug/l
RT: 4.866 min Scan# 517
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 59 Resp: 41440
Ion Ratio Lower Upper
59 100
57 10.8 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#12
1,1-Dichloroethene
Concen: 53.144 ug/l
RT: 3.787 min Scan# 339
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29



Tgt Ion: 96 Resp: 96302
Ion Ratio Lower Upper
96 100
61 167.7 134.9 202.3
98 62.0 49.9 74.9





#13

Acrolein

Concen: 221.842 ug/l

RT: 3.647 min Scan# 317

Delta R.T. -0.006 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 56 Resp: 4585

Ion Ratio Lower Upper

56 100





55 71.9 55.6 83.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025

Supervised By :Semsettin Yesilyurt 05/19/2025



#14

Allyl chloride

Concen: 50.390 ug/l

RT: 4.378 min Scan# 436

Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29




Tgt Ion: 41 Resp: 160107

Ion Ratio Lower Upper

41 100

39 80.8 63.4 95.0


76 39.8 29.5 44.3





#15
Acrylonitrile
Concen: 294.229 ug/l
RT: 5.055 min Scan# 548
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 53 Resp: 122379
Ion Ratio Lower Upper
53 100
52 81.6 66.1 99.1
51 38.0 30.0 45.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#16
Acetone
Concen: 261.381 ug/l
RT: 3.872 min Scan# 353
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29



Tgt Ion: 43 Resp: 90742
Ion Ratio Lower Upper
43 100
58 28.6 23.6 35.4





#17

Carbon Disulfide

Concen: 52.482 ug/l

RT: 4.104 min Scan# 391

Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 310118

Ion Ratio Lower Upper

76 100





78 9.6 7.2 10.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025

Supervised By :Semsettin Yesilyurt 05/19/2025



#18

Methyl Acetate

Concen: 72.098 ug/l

RT: 4.385 min Scan# 437

Delta R.T. -0.000 min

Lab File: VY022301.D




Acq: 16 May 2025 19:29

Tgt Ion: 43 Resp: 78917

Ion Ratio Lower Upper

43 100


74 24.0 18.2 27.4





#19
Methyl tert-butyl Ether
Concen: 57.942 ug/l
RT: 5.110 min Scan# 511
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 73 Resp: 289508
Ion Ratio Lower Upper
73 100
57 21.2 18.6 28.0




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#20
Methylene Chloride
Concen: 50.040 ug/l
RT: 4.610 min Scan# 474
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29


Tgt Ion: 84 Resp: 112252
Ion Ratio Lower Upper
84 100
49 132.5 109.6 164.4
51 40.2 32.4 48.6
86 63.4 52.3 78.5





#21
trans-1,2-Dichloroethene
Concen: 53.696 ug/l
RT: 5.110 min Scan# 511
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 96 Resp: 107400
Ion Ratio Lower Upper
96 100
61 138.9 113.4 170.0
98 63.8 50.9 76.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#22
Diisopropyl ether
Concen: 53.264 ug/l
RT: 6.012 min Scan# 704
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Tgt Ion: 45 Resp: 357394
Ion Ratio Lower Upper
45 100
43 64.1 48.2 72.2
87 27.7 21.4 32.0
59 12.4 9.0 13.4





#23
 Vinyl Acetate
 Concen: 252.608 ug/l
 RT: 5.957 min Scan# 695
 Delta R.T. -0.000 min
 Lab File: VY022301.D
 Acq: 16 May 2025 19:29




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion: 43 Resp: 1023992
 Ion Ratio Lower Upper
 43 100
 86 10.5 7.8 11.8

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
 Supervised By :Semsettin Yesilyurt 05/19/2025




#24
 1,1-Dichloroethane
 Concen: 53.998 ug/l
 RT: 5.909 min Scan# 687
 Delta R.T. -0.006 min
 Lab File: VY022301.D
 Acq: 16 May 2025 19:29



Tgt Ion: 63 Resp: 202421
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.5 10.5
 100 4.4 2.4 7.2









#25
2-Butanone
Concen: 273.834 ug/l
RT: 6.890 min Scan# 849
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 159010
Ion Ratio Lower Upper
43 100
72 27.9 22.1 33.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#26
2,2-Dichloropropane
Concen: 46.353 ug/l
RT: 6.878 min Scan# 846
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29


Tgt Ion: 77 Resp: 154549
Ion Ratio Lower Upper
77 100
97 23.3 11.2 33.5





#27
 cis-1,2-Dichloroethene
 Concen: 55.900 ug/l
 RT: 6.884 min Scan# 848
 Delta R.T. -0.006 min
 Lab File: VY022301.D
 Acq: 16 May 2025 19:29




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion: 96 Resp: 129025
 Ion Ratio Lower Upper
 96 100
 61 145.1 0.0 302.6
 98 64.1 0.0 128.8




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
 Supervised By :Semsettin Yesilyurt 05/19/2025



#28
 Bromochloromethane
 Concen: 54.788 ug/l
 RT: 7.244 min Scan# 906
 Delta R.T. -0.000 min
 Lab File: VY022301.D
 Acq: 16 May 2025 19:29


Tgt Ion: 49 Resp: 87604
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 3.6
 130 71.0 54.7 82.1





#29
Tetrahydrofuran
Concen: 290.562 ug/l
RT: 7.262 min Scan# 909
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 42 Resp: 10855
Ion Ratio Lower Upper
42 100
72 44.1 33.8 50.6
71 42.5 31.5 47.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#30
Chloroform
Concen: 56.301 ug/l
RT: 7.421 min Scan# 935
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Tgt Ion: 83 Resp: 202440
Ion Ratio Lower Upper
83 100
85 67.7 51.8 77.8





#31
Cyclohexane
Concen: 45.799 ug/l
RT: 7.695 min Scan# 981
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 171539
Ion Ratio Lower Upper
56 100
69 31.3 27.4 41.2
84 85.2 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#32
1,1,1-Trichloroethane
Concen: 52.300 ug/l
RT: 7.616 min Scan# 967
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29



Tgt Ion: 97 Resp: 172530
Ion Ratio Lower Upper
97 100
99 64.5 51.4 77.0
61 47.3 38.1 57.1





#33

1,2-Dichloroethane-d4

Concen: 56.783 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 10513

Ion Ratio Lower Upper

65 100





67 52.2 0.0 104.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025

Supervised By :Semsettin Yesilyurt 05/19/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29




Tgt Ion:114 Resp: 301400


Ion Ratio Lower Upper

114 100

63 19.4 0.0 41.0


88 14.6 0.0 29.4





#35
Dibromofluoromethane
Concen: 53.347 ug/l
RT: 7.634 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 113 Resp: 9596
Ion Ratio Lower Upper
113 100
111 102.2 82.6 123.8
192 18.1 15.2 22.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#36
1,1-Dichloropropene
Concen: 48.494 ug/l
RT: 7.835 min Scan# 1003
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Tgt Ion: 75 Resp: 143691
Ion Ratio Lower Upper
75 100
110 33.7 16.6 49.8
77 31.1 24.6 37.0





#37
Ethyl Acetate
Concen: 50.310 ug/l
RT: 6.982 min Scan# 80
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 71720
Ion Ratio Lower Upper
43 100
61 14.2 10.2 15.2
70 9.8 7.6 11.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#38
Carbon Tetrachloride
Concen: 48.210 ug/l
RT: 7.817 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29

Tgt Ion:117 Resp: 147865
Ion Ratio Lower Upper
117 100
119 95.1 75.3 112.9
121 29.9 24.7 37.1





#39

Methylcyclohexane

Concen: 45.224 ug/l

RT: 9.109 min Scan# 1212

Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050EC

Tgt Ion:	83	Resp:	18003
Ion Ratio	Lower	Upper	
83	100		
55	72.6	60.6	90.8
98	48.0	40.0	60.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#40

Benzene

Concen: 51.925 ug/l




RT: 8.079 min Scan# 1043


Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29


Tgt Ion:	78	Resp:	456109
Ion Ratio	Lower	Upper	
78	100		
77	24.6	18.9	28.3





#41
Methacrylonitrile
Concen: 42.134 ug/l
RT: 7.213 min Scan# 901
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 41 Resp: 39840
Ion Ratio Lower Upper
41 100
39 65.5 39.1 58.7
67 77.9 47.1 70.7
52 34.0 21.4 32.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#42
1,2-Dichloroethane
Concen: 53.352 ug/l
RT: 8.152 min Scan# 1055
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29


Tgt Ion: 62 Resp: 126733
Ion Ratio Lower Upper
62 100
98 8.7 0.0 18.4





#43
Isopropyl Acetate
Concen: 51.797 ug/l
RT: 8.195 min Scan# 1062
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 16104
Ion Ratio Lower Upper
43 100
61 26.7 15.1 22.7
87 11.9 9.7 14.5




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#44
Trichloroethene
Concen: 50.616 ug/l
RT: 8.865 min Scan# 1172
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29

Tgt Ion:130 Resp: 110550
Ion Ratio Lower Upper
130 100
95 97.1 0.0 204.8





#45

1,2-Dichloropropane

Concen: 51.337 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 108718

Ion Ratio Lower Upper

63 100





65 30.5 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025

Supervised By :Semsettin Yesilyurt 05/19/2025



#46

Dibromomethane

Concen: 53.928 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29




Tgt Ion: 93 Resp: 61598


Ion Ratio Lower Upper

93 100

95 82.2 65.2 97.8


174 90.7 74.3 111.5





#47
Bromodichloromethane
Concen: 52.627 ug/l
RT: 9.420 min Scan# 1263
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 155944
Ion Ratio Lower Upper
83 100
85 61.8 51.1 76.7
127 7.4 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#48
Methyl methacrylate
Concen: 51.751 ug/l
RT: 9.219 min Scan# 1230
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29


Tgt Ion: 41 Resp: 75248
Ion Ratio Lower Upper
41 100
69 84.7 63.5 95.3
39 67.9 50.8 76.2





#49
1,4-Dioxane
Concen: 1116.500 ug/l
RT: 9.237 min Scan# 1232
Delta R.T. 0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 88 Resp: 1440
Ion Ratio Lower Upper
88 100
43 34.0 28.5 42.7
58 60.6 61.4 92.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#50
Toluene-d8
Concen: 52.262 ug/l
RT: 10.103 min Scan# 1375
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29


Tgt Ion: 98 Resp: 385048
Ion Ratio Lower Upper
98 100
100 64.8 51.8 77.8





#51
4-Methyl-2-Pentanone
Concen: 270.195 ug/l
RT: 9.993 min Scan# 1358
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 41221
Ion Ratio Lower Upper
43 100
58 38.2 29.4 44.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#52
Toluene
Concen: 51.150 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29


Tgt Ion: 92 Resp: 284515
Ion Ratio Lower Upper
92 100
91 175.1 139.1 208.7





#53
t-1,3-Dichloropropene
Concen: 50.870 ug/l
RT: 10.390 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 75 Resp: 147108
Ion Ratio Lower Upper
75 100
77 30.5 25.0 37.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#54
cis-1,3-Dichloropropene
Concen: 51.085 ug/l
RT: 9.853 min Scan# 1334
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Tgt Ion: 75 Resp: 169256
Ion Ratio Lower Upper
75 100
77 31.3 25.1 37.7
39 53.2 44.5 66.7





#55
1,1,2-Trichloroethane
Concen: 53.400 ug/l
RT: 10.572 min Scan# 1452
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 97 Resp: 7810
Ion Ratio Lower Upper
97 100
83 90.3 70.3 105.5
85 59.1 46.2 69.4
99 64.3 50.3 75.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#56
Ethyl methacrylate
Concen: 52.876 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Tgt Ion: 69 Resp: 120471
Ion Ratio Lower Upper
69 100
41 69.2 56.5 84.7
39 46.8 37.1 55.7





#57
1,3-Dichloropropane
Concen: 53.333 ug/l
RT: 10.713 min Scan# 1476
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 76 Resp: 13860
Ion Ratio Lower Upper
76 100
78 32.9 25.8 38.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#58
2-Chloroethyl Vinyl ether
Concen: 294.015 ug/l
RT: 9.707 min Scan# 1310
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Tgt Ion: 63 Resp: 300672
Ion Ratio Lower Upper
63 100
106 26.4 21.0 31.4





#59
2-Hexanone
Concen: 264.900 ug/l
RT: 10.755 min Scan# 1482
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 270964
Ion Ratio Lower Upper
43 100
58 53.9 26.1 78.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#60
Dibromochloromethane
Concen: 53.353 ug/l
RT: 10.908 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29


Tgt Ion:129 Resp: 101289
Ion Ratio Lower Upper
129 100
127 77.8 38.6 115.7





#61
1,2-Dibromoethane
Concen: 54.113 ug/l
RT: 11.011 min Scan# 1524
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:107 Resp: 73769
Ion Ratio Lower Upper
107 100
109 93.4 74.6 111.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#62
4-Bromofluorobenzene
Concen: 50.802 ug/l
RT: 12.401 min Scan# 1752
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29



Tgt Ion: 95 Resp: 118638
Ion Ratio Lower Upper
95 100
174 84.1 0.0 166.8
176 81.9 0.0 160.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC





Tgt Ion:117 Resp: 26298

Ion	Ratio	Lower	Upper
117	100		
82	56.8	46.6	70.0
119	31.2	25.8	38.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#64

Tetrachloroethene

Concen: 48.141 ug/l

RT: 10.646 min Scan# 1464




Delta R.T. -0.000 min


Lab File: VY022301.D

Acq: 16 May 2025 19:29

Tgt Ion:164 Resp: 117020

Ion	Ratio	Lower	Upper
164	100		
166	123.1	100.4	150.6
129	90.1	76.4	114.6
131	86.2	73.1	109.7









#65
Chlorobenzene
Concen: 50.288 ug/l
RT: 11.438 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:112 Resp: 29836
Ion Ratio Lower Upper
112 100
114 33.5 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 50.525 ug/l
RT: 11.511 min Scan# 1606
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29


Tgt Ion:131 Resp: 103057
Ion Ratio Lower Upper
131 100
133 96.3 47.3 141.9
119 66.8 33.1 99.2





#67
Ethyl Benzene
Concen: 49.681 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 55203
Ion Ratio Lower Upper
91 100
106 30.4 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#68
m/p-Xylenes
Concen: 99.542 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Tgt Ion:106 Resp: 416782
Ion Ratio Lower Upper
106 100
91 202.8 163.0 244.4





#69
o-Xylene
Concen: 50.639 ug/l
RT: 11.950 min Scan# 1678
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:106 Resp: 200603
Ion Ratio Lower Upper
106 100
91 214.6 108.3 324.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#70
Styrene
Concen: 51.248 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29



Tgt Ion:104 Resp: 342456
Ion Ratio Lower Upper
104 100
78 49.2 40.9 61.3
103 54.9 43.7 65.5





#71

Bromoform

Concen: 52.719 ug/l

RT: 12.133 min Scan# 1708

Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050EC

Tgt Ion:173 Resp: 5804
 Ion Ratio Lower Upper
 173 100
 175 48.6 24.7 74.1
 254 0.2 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025
 Supervised By :Semsettin Yesilyurt 05/19/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.346 min Scan# 1907


Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29


Tgt Ion:152 Resp: 124169
 Ion Ratio Lower Upper
 152 100
 115 58.0 29.4 88.2
 150 176.7 0.0 353.8





#73
Isopropylbenzene
Concen: 49.232 ug/l
RT: 12.249 min Scan# 1727
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:105 Resp: 50644
Ion Ratio Lower Upper
105 100
120 26.3 13.2 39.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#74
N-ethyl acetate
Concen: 49.313 ug/l
RT: 12.066 min Scan# 1697
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Tgt Ion: 43 Resp: 145082
Ion Ratio Lower Upper
43 100
70 40.1 31.4 47.0
55 21.2 17.0 25.6
61 23.1 18.6 28.0





#75
1,1,2,2-Tetrachloroethane
Concen: 54.429 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 83 Resp: 86108
Ion Ratio Lower Upper
83 100
131 9.4 4.9 14.7
85 63.2 32.0 96.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#76
1,2,3-Trichloropropane
Concen: 49.451 ug/l m
RT: 12.554 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29



Tgt Ion: 75 Resp: 62208
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 50.903 ug/l

RT: 12.529 min Scan# 1773

Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050EC

Tgt Ion:	156	Resp:	11183
Ion Ratio	Lower	Upper	
156	100		
77	190.0	94.5	283.6
158	96.8	48.4	145.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#78

n-propylbenzene

Concen: 49.582 ug/l




RT: 12.590 min Scan# 1783


Delta R.T. -0.006 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Tgt Ion:	91	Resp:	615206
Ion Ratio	Lower	Upper	
91	100		
120	21.8	11.1	33.3









#79
2-Chlorotoluene
Concen: 49.744 ug/l
RT: 12.676 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 33918
Ion Ratio Lower Upper
91 100
126 33.7 16.8 50.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#80
1,3,5-Trimethylbenzene
Concen: 48.863 ug/l
RT: 12.731 min Scan# 1806
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29


Tgt Ion:105 Resp: 405753
Ion Ratio Lower Upper
105 100
120 49.0 24.2 72.6





#81
trans-1,4-Dichloro-2-butene
Concen: 46.262 ug/l
RT: 12.298 min Scan# 1735
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 28728
Ion Ratio Lower Upper
75 100
53 94.4 75.8 113.8
89 45.1 36.2 54.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#82
4-Chlorotoluene
Concen: 49.230 ug/l
RT: 12.773 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29

Tgt Ion: 91 Resp: 345271
Ion Ratio Lower Upper
91 100
126 34.5 16.8 50.4





#83

tert-Butylbenzene

Concen: 50.023 ug/l

RT: 12.993 min Scan# 1849

Delta R.T. -0.006 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion:119 Resp: 369198





Ion	Ratio	Lower	Upper
119	100		
91	64.0	32.8	98.3
134	25.4	12.9	38.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025

Supervised By :Semsettin Yesilyurt 05/19/2025



#84

1,2,4-Trimethylbenzene

Concen: 49.978 ug/l

RT: 13.042 min Scan# 1857




Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Tgt Ion:105 Resp: 411972


Ion	Ratio	Lower	Upper
105	100		
120	45.6	22.3	66.9





#85
 sec-Butylbenzene
 Concen: 49.149 ug/l
 RT: 13.176 min Scan# 1879
 Delta R.T. -0.000 min
 Lab File: VY022301.D
 Acq: 16 May 2025 19:29




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion:105 Resp: 538470
 Ion Ratio Lower Upper
 105 100
 134 19.7 9.8 29.5

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
 Supervised By :Semsettin Yesilyurt 05/19/2025




#86
 p-Isopropyltoluene
 Concen: 48.814 ug/l
 RT: 13.291 min Scan# 1898
 Delta R.T. -0.000 min
 Lab File: VY022301.D
 Acq: 16 May 2025 19:29



Tgt Ion:119 Resp: 449798
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.9 38.6
 91 23.7 12.1 36.3






#87
 1,3-Dichlorobenzene
 Concen: 50.006 ug/l
 RT: 13.285 min Scan# 1897
 Delta R.T. -0.000 min
 Lab File: VY022301.D
 Acq: 16 May 2025 19:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC






Tgt Ion:146 Resp: 224283
 Ion Ratio Lower Upper
 146 100
 111 38.9 19.9 59.7
 148 64.2 31.9 95.7


Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
 Supervised By :Semsettin Yesilyurt 05/19/2025




#88
 1,4-Dichlorobenzene
 Concen: 50.503 ug/l
 RT: 13.365 min Scan# 1910
 Delta R.T. -0.000 min
 Lab File: VY022301.D
 Acq: 16 May 2025 19:29




Tgt Ion:146 Resp: 217882
 Ion Ratio Lower Upper
 146 100
 111 37.1 20.0 59.9
 148 63.6 32.1 96.5





#89
n-Butylbenzene
Concen: 48.542 ug/l
RT: 13.615 min Scan# 1951
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 42315
Ion Ratio Lower Upper
91 100
92 52.0 26.3 78.9
134 24.3 12.4 37.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#90
Hexachloroethane
Concen: 48.950 ug/l
RT: 13.877 min Scan# 1994
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Tgt Ion:117 Resp: 91050
Ion Ratio Lower Upper
117 100
201 69.7 35.2 105.6





#91
1,2-Dichlorobenzene
Concen: 51.391 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 193609
Ion Ratio Lower Upper
146 100
111 40.0 20.5 61.5
148 62.9 32.0 96.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 52.958 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29



Tgt Ion: 75 Resp: 13510
Ion Ratio Lower Upper
75 100
155 85.8 43.4 130.1
157 108.2 54.6 163.9





#93

1,2,4-Trichlorobenzene

Concen: 49.489 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050EC

Tgt Ion:180	Resp: 10769
Ion Ratio	Lower Upper
180	100
182	95.6 48.0 144.0
145	29.9 15.0 45.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#94

Hexachlorobutadiene

Concen: 46.635 ug/l




RT: 15.023 min Scan# 2182


Delta R.T. -0.000 min

Lab File: VY022301.D

Acq: 16 May 2025 19:29


Tgt Ion:225	Resp: 56596
Ion Ratio	Lower Upper
225	100
223	60.2 31.5 94.5
227	63.1 32.1 96.3





#95
Naphthalene
Concen: 52.582 ug/l
RT: 15.139 min Scan# 217144
Delta R.T. -0.006 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:128 Resp: 217144
Ion Ratio Lower Upper
128 100
127 12.9 10.2 15.2
129 11.0 8.8 13.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/19/2025
Supervised By :Semsettin Yesilyurt 05/19/2025



#96
1,2,3-Trichlorobenzene
Concen: 50.540 ug/l
RT: 15.328 min Scan# 2232
Delta R.T. -0.000 min
Lab File: VY022301.D
Acq: 16 May 2025 19:29

Tgt Ion:180 Resp: 93303
Ion Ratio Lower Upper
180 100
182 94.8 47.1 141.3
145 31.1 15.9 47.6

