

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051722\
 Data File : VY008780.D
 Acq On : 17 May 2022 18:44
 Operator : KP/MD
 Sample : VY0517SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0517SBL01

Quant Time: May 18 07:56:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 07:32:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	119303	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	202586	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	179356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	74973	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	64786	58.536	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.080%
35) Dibromofluoromethane	7.716	113	63237	53.425	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.840%
50) Toluene-d8	10.179	98	245643	46.898	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.800%
62) 4-Bromofluorobenzene	12.483	95	68226	37.289	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	74.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051722\
Data File : VY008780.D
Acq On : 17 May 2022 18:44
Operator : KP/MD
Sample : VY0517SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0517SBL01

Quant Time: May 18 07:56:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 07:32:50 2022
Response via : Initial Calibration

