

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051722\
 Data File : VY008804.D
 Acq On : 18 May 2022 04:27
 Operator : KP/MD
 Sample : N2869-03
 Misc : 4.27g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P027-SS002-0006-01

Quant Time: May 18 08:02:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 07:32:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	110938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	196556	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	184651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	80567	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	69253	67.290	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.580%
35) Dibromofluoromethane	7.716	113	65719	57.225	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.440%
50) Toluene-d8	10.179	98	243577	47.861	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.720%
62) 4-Bromofluorobenzene	12.483	95	75896	42.243	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.480%
Target Compounds						
11) Tert butyl alcohol	4.948	59	6525	57.365	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051722\
Data File : VY008804.D
Acq On : 18 May 2022 04:27
Operator : KP/MD
Sample : N2869-03
Misc : 4.27g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P027-SS002-0006-01

Quant Time: May 18 08:02:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 07:32:50 2022
Response via : Initial Calibration

