

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051922\
 Data File : VY008880.D
 Acq On : 19 May 2022 20:05
 Operator : KP/MD
 Sample : N2868-16
 Misc : 5.44g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P022-SS002-0006-01

Quant Time: May 20 02:54:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 07:32:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	129449	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	236897	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	224990	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	95885	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	83904	69.867	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	139.740%
35) Dibromofluoromethane	7.716	113	77247	55.809	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.620%
50) Toluene-d8	10.179	98	299772	48.805	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.600%
62) 4-Bromofluorobenzene	12.483	95	85545	39.732	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	79.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051922\
Data File : VY008880.D
Acq On : 19 May 2022 20:05
Operator : KP/MD
Sample : N2868-16
Misc : 5.44g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P022-SS002-0006-01

Quant Time: May 20 02:54:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 07:32:50 2022
Response via : Initial Calibration

