

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051922\
 Data File : VY008894.D
 Acq On : 20 May 2022 02:38
 Operator : KP/MD
 Sample : N2907-03
 Misc : 5.47g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GC-3

Quant Time: May 20 04:29:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 07:32:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	121514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	212057	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	197880	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	84079	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	71264	63.217	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.440%
35) Dibromofluoromethane	7.716	113	66802	53.916	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.840%
50) Toluene-d8	10.179	98	267218	48.614	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.220%
62) 4-Bromofluorobenzene	12.483	95	74259	38.635	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051922\
Data File : VY008894.D
Acq On : 20 May 2022 02:38
Operator : KP/MD
Sample : N2907-03
Misc : 5.47g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
GC-3

Quant Time: May 20 04:29:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 07:32:50 2022
Response via : Initial Calibration

