

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051922\
 Data File : VY008883.D
 Acq On : 19 May 2022 21:13
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/24/2022
 Supervised By : Mahesh Dadoda 05/24/2022

Quant Time: May 20 02:55:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 07:32:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	140303	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	239344	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	249980	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	153750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	70033	53.805	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	107.620%		
35) Dibromofluoromethane	7.716	113	67516	48.280	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	96.560%		
50) Toluene-d8	10.179	98	283035	45.817	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	91.640%		
62) 4-Bromofluorobenzene	12.483	95	107492	48.565	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	97.140%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.906	85	44814	51.539	ug/l	99
3) Chloromethane	2.113	50	72893	59.387	ug/l	97
4) Vinyl Chloride	2.247	62	104353	55.765	ug/l	97
5) Bromomethane	2.649	94	93128	57.198	ug/l	99
6) Chloroethane	2.790	64	77840	54.167	ug/l	98
7) Trichlorofluoromethane	3.125	101	95172	50.687	ug/l	92
8) Diethyl Ether	3.527	74	34102	54.857	ug/l	81
9) 1,1,2-Trichlorotrifluo...	3.893	101	63004	54.959	ug/l	92
10) Methyl Iodide	4.088	142	70782	48.695	ug/l	91
11) Tert butyl alcohol	4.954	59	30775	213.933	ug/l #	92
12) 1,1-Dichloroethene	3.869	96	55360	46.501	ug/l	93
13) Acrolein	3.729	56	7558	106.206	ug/l	95
14) Allyl chloride	4.472	41	91434	52.213	ug/l #	89
15) Acrylonitrile	5.161	53	98911	234.637	ug/l	97
16) Acetone	3.948	43	68279	233.355	ug/l	90
17) Carbon Disulfide	4.192	76	155317	50.931	ug/l	99
18) Methyl Acetate	4.478	43	43848	47.165	ug/l #	89
19) Methyl tert-butyl Ether	5.216	73	192030	59.891	ug/l	96
20) Methylene Chloride	4.716	84	73910	55.864	ug/l	84
21) trans-1,2-Dichloroethene	5.216	96	74585	51.095	ug/l	88
22) Diisopropyl ether	6.118	45	242685	65.439	ug/l #	94
23) Vinyl Acetate	6.057	43	784992	300.209	ug/l #	92
24) 1,1-Dichloroethane	6.015	63	138061	56.104	ug/l	98
25) 2-Butanone	6.984	43	128679	235.577	ug/l #	88
26) 2,2-Dichloropropane	6.978	77	117664	50.308	ug/l	95
27) cis-1,2-Dichloroethene	6.984	96	92257	56.982	ug/l	88
28) Bromochloromethane	7.332	49	49123	52.654	ug/l #	88
29) Tetrahydrofuran	7.344	42	86754	238.418	ug/l #	85
30) Chloroform	7.508	83	151581	57.453	ug/l	98
31) Cyclohexane	7.783	56	114025	55.819	ug/l	90
32) 1,1,1-Trichloroethane	7.703	97	133780	54.690	ug/l	97
36) 1,1-Dichloropropene	7.917	75	109992	49.109	ug/l	99
37) Ethyl Acetate	7.076	43	59362	44.475	ug/l	97
38) Carbon Tetrachloride	7.899	117	118727	49.462	ug/l	98
39) Methylcyclohexane	9.185	83	137513	51.772	ug/l #	88
40) Benzene	8.161	78	341137	54.433	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051922\
 Data File : VY008883.D
 Acq On : 19 May 2022 21:13
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/24/2022
 Supervised By : Mahesh Dadoda 05/24/2022

Quant Time: May 20 02:55:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 07:32:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.325	41	37711m	53.725	ug/l	
42) 1,2-Dichloroethane	8.234	62	104322	56.708	ug/l	92
43) Isopropyl Acetate	8.270	43	118417	53.295	ug/l #	92
44) Trichloroethene	8.941	130	88677	52.016	ug/l	99
45) 1,2-Dichloropropane	9.215	63	86745	58.240	ug/l	95
46) Dibromomethane	9.307	93	53916	55.089	ug/l	94
47) Bromodichloromethane	9.496	83	128370	59.165	ug/l	96
48) Methyl methacrylate	9.289	41	53171	54.499	ug/l #	82
49) 1,4-Dioxane	9.295	88	13941	948.328	ug/l #	83
51) 4-Methyl-2-Pentanone	10.069	43	337728	257.474	ug/l	91
52) Toluene	10.246	92	243649	57.214	ug/l	95
53) t-1,3-Dichloropropene	10.465	75	131892	58.524	ug/l	99
54) cis-1,3-Dichloropropene	9.929	75	141274	58.132	ug/l #	87
55) 1,1,2-Trichloroethane	10.642	97	79168	59.485	ug/l	98
56) Ethyl methacrylate	10.508	69	101547	62.888	ug/l #	82
57) 1,3-Dichloropropane	10.788	76	130104	59.452	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.782	63	226145	274.648	ug/l	94
59) 2-Hexanone	10.831	43	230330	245.759	ug/l	87
60) Dibromochloromethane	10.983	129	97360	60.937	ug/l	99
61) 1,2-Dibromoethane	11.087	107	75264	56.707	ug/l	100
64) Tetrachloroethene	10.721	164	78624	49.135	ug/l	99
65) Chlorobenzene	11.514	112	267840	52.940	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.587	131	105820	57.251	ug/l	96
67) Ethyl Benzene	11.593	91	491700	54.421	ug/l	98
68) m/p-Xylenes	11.703	106	407016	113.078	ug/l	91
69) o-Xylene	12.032	106	194540	58.494	ug/l	90
70) Styrene	12.044	104	344857	61.253	ug/l	93
71) Bromoform	12.209	173	69585	55.458	ug/l #	98
73) Isopropylbenzene	12.331	105	526323	49.913	ug/l	100
74) N-amyl acetate	12.142	43	127866	50.802	ug/l #	85
75) 1,1,2,2-Tetrachloroethane	12.581	83	106590	45.469	ug/l	98
76) 1,2,3-Trichloropropane	12.635	75	80431m	45.287	ug/l	
77) Bromobenzene	12.611	156	121433	49.700	ug/l	93
78) n-propylbenzene	12.672	91	648876	50.090	ug/l	99
79) 2-Chlorotoluene	12.757	91	361719	51.553	ug/l	97
80) 1,3,5-Trimethylbenzene	12.812	105	459962	54.025	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.379	75	36272	44.671	ug/l #	85
82) 4-Chlorotoluene	12.855	91	399480	53.893	ug/l	97
83) tert-Butylbenzene	13.074	119	408900	53.404	ug/l	94
84) 1,2,4-Trimethylbenzene	13.123	105	464219	55.078	ug/l	97
85) sec-Butylbenzene	13.257	105	635772	53.221	ug/l	100
86) p-Isopropyltoluene	13.373	119	532059	54.623	ug/l	97
87) 1,3-Dichlorobenzene	13.367	146	275046	53.575	ug/l	100
88) 1,4-Dichlorobenzene	13.446	146	273340	53.828	ug/l	99
89) n-Butylbenzene	13.696	91	476224	51.508	ug/l	100
90) Hexachloroethane	13.958	117	97912	53.166	ug/l	84
91) 1,2-Dichlorobenzene	13.739	146	245200	54.616	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.361	75	15754	39.304	ug/l	75
93) 1,2,4-Trichlorobenzene	15.007	180	128162	52.276	ug/l	99
94) Hexachlorobutadiene	15.117	225	68701	48.698	ug/l	98
95) Naphthalene	15.239	128	267292	48.770	ug/l	99
96) 1,2,3-Trichlorobenzene	15.428	180	113561	52.729	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051922\
Data File : VY008883.D
Acq On : 19 May 2022 21:13
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022

Quant Time: May 20 02:55:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 07:32:50 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051922\
 Data File : VY008883.D
 Acq On : 19 May 2022 21:13
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/soil
 ALS Vial : 28 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Quant Time: May 20 02:55:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 07:32:50 2022
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By : John Carlone 05/24/2022
 Supervised By : Mahesh Dadoda 05/24/2022





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 14030
Ion Ratio Lower Upper
168 100
99 54.8 46.9 70.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#2
Dichlorodifluoromethane
Concen: 51.539 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion: 85 Resp: 44814
Ion Ratio Lower Upper
85 100
87 32.5 16.1 48.2





#3
Chloromethane
Concen: 59.387 ug/l
RT: 2.113 min Scan# 41
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 50 Resp: 7289
Ion Ratio Lower Upper
50 100
52 32.3 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#4
Vinyl Chloride
Concen: 55.765 ug/l
RT: 2.247 min Scan# 70
Delta R.T. 0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion: 62 Resp: 104353
Ion Ratio Lower Upper
62 100
64 32.9 25.0 37.6





#5
Bromomethane
Concen: 57.198 ug/l
RT: 2.649 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 93128
Ion Ratio Lower Upper
94 100
96 91.4 73.5 110.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#6
Chloroethane
Concen: 54.167 ug/l
RT: 2.790 min Scan# 159
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion: 64 Resp: 77840
Ion Ratio Lower Upper
64 100
66 32.7 25.4 38.0





#7
Trichlorofluoromethane
Concen: 50.687 ug/l
RT: 3.125 min Scan# 213
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:101 Resp: 95172
Ion Ratio Lower Upper
101 100
103 66.1 48.0 72.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#8
Diethyl Ether
Concen: 54.857 ug/l
RT: 3.527 min Scan# 280
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13



Tgt Ion: 74 Resp: 34102
Ion Ratio Lower Upper
74 100
45 90.6 55.1 165.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 54.959 ug/l
RT: 3.893 min Scan# 340
Delta R.T. 0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 63004
Ion Ratio Lower Upper
101 100
85 45.7 40.6 60.8
151 73.9 64.9 97.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#10
Methyl Iodide
Concen: 48.695 ug/l
RT: 4.088 min Scan# 372
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13

Tgt Ion:142 Resp: 70782
Ion Ratio Lower Upper
142 100
127 44.3 41.5 62.3
141 14.1 11.9 17.9





#11

Tert butyl alcohol

Concen: 213.933 ug/l

RT: 4.954 min Scan# 512

Delta R.T. 0.012 min

Lab File: VY008883.D


Acq: 19 May 2022 21:13

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 59 Resp: 30775

Ion Ratio Lower Upper

59 100



57 10.7 6.4 9.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/24/2022

Supervised By :Mahesh Dadoda 05/24/2022



#12

1,1-Dichloroethene

Concen: 46.501 ug/l

RT: 3.869 min Scan# 336

Delta R.T. -0.000 min

Lab File: VY008883.D

Acq: 19 May 2022 21:13




Tgt Ion: 96 Resp: 55360


Ion Ratio Lower Upper

96 100

61 167.7 145.8 218.6


98 65.4 52.5 78.7





#13
Acrolein
Concen: 106.206 ug/l
RT: 3.729 min Scan# 313
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 7558
Ion Ratio Lower Upper
56 100
55 67.9 57.6 86.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#14
Allyl chloride
Concen: 52.213 ug/l
RT: 4.472 min Scan# 435
Delta R.T. -0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion: 41 Resp: 91434
Ion Ratio Lower Upper
41 100
39 79.7 58.8 88.2
76 41.8 25.1 37.7#





#15
Acrylonitrile
Concen: 234.637 ug/l
RT: 5.161 min Scan# 547
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 9891
Ion Ratio Lower Upper
53 100
52 80.3 67.0 100.4
51 36.5 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#16
Acetone
Concen: 233.355 ug/l
RT: 3.948 min Scan# 349
Delta R.T. 0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion: 43 Resp: 68279
Ion Ratio Lower Upper
43 100
58 32.7 22.0 33.0





#17
Carbon Disulfide
Concen: 50.931 ug/l
RT: 4.192 min Scan# 389
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 76 Resp: 155317
Ion Ratio Lower Upper
76 100
78 8.8 7.4 11.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#18
Methyl Acetate
Concen: 47.165 ug/l
RT: 4.478 min Scan# 436
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion: 43 Resp: 43848
Ion Ratio Lower Upper
43 100
74 25.8 16.6 24.8





#19
Methyl tert-butyl Ether
Concen: 59.891 ug/l
RT: 5.216 min Scan# 51
Delta R.T. -0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 73 Resp: 192030
Ion Ratio Lower Upper
73 100
57 21.2 18.6 27.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#20
Methylene Chloride
Concen: 55.864 ug/l
RT: 4.716 min Scan# 475
Delta R.T. 0.012 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion: 84 Resp: 73910
Ion Ratio Lower Upper
84 100
49 116.3 114.4 171.6
51 36.5 35.3 52.9
86 60.8 53.7 80.5





#21
trans-1,2-Dichloroethene
Concen: 51.095 ug/l
RT: 5.216 min Scan# 51
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 96 Resp: 7458
Ion Ratio Lower Upper
96 100
61 130.7 120.5 180.7
98 63.7 49.8 74.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#22
Diisopropyl ether
Concen: 65.439 ug/l
RT: 6.118 min Scan# 705
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion: 45 Resp: 242685
Ion Ratio Lower Upper
45 100
43 57.8 49.3 73.9
87 29.1 19.2 28.8#
59 12.6 8.4 12.6





#23
 Vinyl Acetate
 Concen: 300.209 ug/l
 RT: 6.057 min Scan# 695
 Delta R.T. -0.000 min
 Lab File: VY008883.D
 Acq: 19 May 2022 21:13




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion: 43 Resp: 784992
 Ion Ratio Lower Upper
 43 100
 86 11.9 7.1 10.7

Manual Integrations
 APPROVED



Reviewed By :John Carlone 05/24/2022
 Supervised By :Mahesh Dadoda 05/24/2022




#24
 1,1-Dichloroethane
 Concen: 56.104 ug/l
 RT: 6.015 min Scan# 688
 Delta R.T. -0.000 min
 Lab File: VY008883.D
 Acq: 19 May 2022 21:13



Tgt Ion: 63 Resp: 138061
 Ion Ratio Lower Upper
 63 100
 98 7.8 3.6 10.8
 100 5.1 2.0 5.8





#25

2-Butanone

Concen: 235.577 ug/l

RT: 6.984 min Scan# 847

Delta R.T. -0.000 min

Lab File: VY008883.D

Acq: 19 May 2022 21:13

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 128679

Ion Ratio Lower Upper

43 100





72 27.8 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/24/2022

Supervised By :Mahesh Dadoda 05/24/2022



#26

2,2-Dichloropropane

Concen: 50.308 ug/l

RT: 6.978 min Scan# 846

Delta R.T. -0.000 min

Lab File: VY008883.D




Acq: 19 May 2022 21:13


Tgt Ion: 77 Resp: 117664

Ion Ratio Lower Upper

77 100


97 23.1 10.4 31.4





#27
cis-1,2-Dichloroethene
Concen: 56.982 ug/l
RT: 6.984 min Scan# 847
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 9225
Ion Ratio Lower Upper
96 100
61 137.4 0.0 319.8
98 64.2 0.0 127.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#28
Bromochloromethane
Concen: 52.654 ug/l
RT: 7.332 min Scan# 904
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion: 49 Resp: 49123
Ion Ratio Lower Upper
49 100
129 2.5 0.0 2.2#
130 79.3 55.4 83.0





#29
Tetrahydrofuran
Concen: 238.418 ug/l
RT: 7.344 min Scan# 907
Delta R.T. -0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 42 Resp: 86754
Ion Ratio Lower Upper
42 100
72 47.2 30.3 45.5
71 44.7 29.0 43.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#30
Chloroform
Concen: 57.453 ug/l
RT: 7.508 min Scan# 933
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion: 83 Resp: 151581
Ion Ratio Lower Upper
83 100
85 66.1 51.8 77.6





#31
Cyclohexane
Concen: 55.819 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 11402
Ion Ratio Lower Upper
56 100
69 31.7 26.7 40.1
84 96.0 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#32
1,1,1-Trichloroethane
Concen: 54.690 ug/l
RT: 7.703 min Scan# 965
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13



Tgt Ion: 97 Resp: 133780
Ion Ratio Lower Upper
97 100
99 64.6 50.7 76.1
61 44.5 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 53.805 ug/l

RT: 8.142 min Scan# 1037

Delta R.T. -0.000 min

Lab File: VY008883.D

Acq: 19 May 2022 21:13

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 7003

Ion Ratio Lower Upper

65 100





67 54.9 0.0 99.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/24/2022

Supervised By :Mahesh Dadoda 05/24/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.000 min

Lab File: VY008883.D

Acq: 19 May 2022 21:13




Tgt Ion:114 Resp: 239344


Ion Ratio Lower Upper

114 100

63 19.1 0.0 43.2


88 13.5 0.0 31.8





#35
Dibromofluoromethane
Concen: 48.280 ug/l
RT: 7.716 min Scan# 90
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 113 Resp: 67510
Ion Ratio Lower Upper
113 100
111 102.6 82.6 124.0
192 19.6 17.6 26.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#36
1,1-Dichloropropene
Concen: 49.109 ug/l
RT: 7.917 min Scan# 1000
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion: 75 Resp: 109992
Ion Ratio Lower Upper
75 100
110 35.6 17.1 51.2
77 30.5 24.4 36.6





#37
Ethyl Acetate
Concen: 44.475 ug/l
RT: 7.076 min Scan# 861
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 59362
Ion Ratio Lower Upper
43 100
61 13.7 9.7 14.5
70 9.4 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#38
Carbon Tetrachloride
Concen: 49.462 ug/l
RT: 7.899 min Scan# 997
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion:117 Resp: 118727
Ion Ratio Lower Upper
117 100
119 95.8 77.9 116.9
121 33.3 24.8 37.2





#39
Methylcyclohexane
Concen: 51.772 ug/l
RT: 9.185 min Scan# 1207
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 13751
Ion Ratio Lower Upper
83 100
55 66.2 66.5 99.7
98 46.6 37.9 56.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#40
Benzene
Concen: 54.433 ug/l
RT: 8.161 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion: 78 Resp: 341137
Ion Ratio Lower Upper
78 100
77 24.5 20.4 30.6





#41
Methacrylonitrile
Concen: 53.725 ug/l m
RT: 7.325 min Scan# 902
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 41 Resp: 3771
Ion Ratio Lower Upper
41 100
39 110.9 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#42
1,2-Dichloroethane
Concen: 56.708 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion: 62 Resp: 104322
Ion Ratio Lower Upper
62 100
98 10.1 0.0 14.8





#43
Isopropyl Acetate
Concen: 53.295 ug/l
RT: 8.270 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 11841
Ion Ratio Lower Upper
43 100
61 29.1 20.2 30.2
87 14.0 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#44
Trichloroethene
Concen: 52.016 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13

Tgt Ion:130 Resp: 88677
Ion Ratio Lower Upper
130 100
95 95.5 0.0 188.2





#45

1,2-Dichloropropane

Concen: 58.240 ug/l

RT: 9.215 min Scan# 1213

Delta R.T. -0.000 min

Lab File: VY008883.D

Acq: 19 May 2022 21:13

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 8674

Ion Ratio Lower Upper

63 100





65 31.6 23.0 34.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/24/2022

Supervised By :Mahesh Dadoda 05/24/2022



#46

Dibromomethane

Concen: 55.089 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. -0.000 min

Lab File: VY008883.D

Acq: 19 May 2022 21:13




Tgt Ion: 93 Resp: 53916


Ion Ratio Lower Upper

93 100

95 82.8 64.3 96.5


174 94.8 83.4 125.0





#47
Bromodichloromethane
Concen: 59.165 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 83 Resp: 128370
Ion Ratio Lower Upper
83 100
85 61.9 51.9 77.9
127 7.8 6.6 10.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#48
Methyl methacrylate
Concen: 54.499 ug/l
RT: 9.289 min Scan# 1225
Delta R.T. -0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion: 41 Resp: 53171
Ion Ratio Lower Upper
41 100
69 94.1 58.2 87.2#
39 65.0 47.4 71.0





#49
1,4-Dioxane
Concen: 948.328 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 88 Resp: 1394
Ion Ratio Lower Upper
88 100
43 31.7 32.8 49.2
58 59.0 59.8 89.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#50
Toluene-d8
Concen: 45.817 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion: 98 Resp: 283035
Ion Ratio Lower Upper
98 100
100 64.6 50.2 75.2





#51
4-Methyl-2-Pentanone
Concen: 257.474 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 337728
Ion Ratio Lower Upper
43 100
58 40.1 28.1 42.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#52
Toluene
Concen: 57.214 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13



Tgt Ion: 92 Resp: 243649
Ion Ratio Lower Upper
92 100
91 171.2 142.6 213.8









#53
t-1,3-Dichloropropene
Concen: 58.524 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13

Instrument : MSVOA_Y
ClientSampleId : VSTDCCC050EC

Tgt Ion: 75 Resp: 131892
Ion Ratio Lower Upper
75 100
77 32.6 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#54
cis-1,3-Dichloropropene
Concen: 58.132 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion: 75 Resp: 141274
Ion Ratio Lower Upper
75 100
77 31.7 25.6 38.4
39 45.1 47.6 71.4#





#55
1,1,2-Trichloroethane
Concen: 59.485 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 97 Resp: 7916


Ion	Ratio	Lower	Upper
97	100		
83	90.0	69.8	104.8
85	55.3	44.3	66.5
99	63.9	49.8	74.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022






#56
Ethyl methacrylate
Concen: 62.888 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13



Tgt Ion: 69 Resp: 101547

Ion	Ratio	Lower	Upper
69	100		
41	62.2	63.4	95.0#
39	35.4	35.8	53.8#





#57

1,3-Dichloropropane

Concen: 59.452 ug/l

RT: 10.788 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VY008883.D

Acq: 19 May 2022 21:13

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 130104

Ion Ratio Lower Upper

76 100





78 31.9 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/24/2022

Supervised By :Mahesh Dadoda 05/24/2022



#58

2-Chloroethyl Vinyl ether

Concen: 274.648 ug/l

RT: 9.782 min Scan# 1306

Delta R.T. -0.000 min

Lab File: VY008883.D




Acq: 19 May 2022 21:13


Tgt Ion: 63 Resp: 226145

Ion Ratio Lower Upper

63 100


106 29.8 21.5 32.3





#59
2-Hexanone
Concen: 245.759 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 230330
Ion Ratio Lower Upper
43 100
58 57.7 24.4 73.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#60
Dibromochloromethane
Concen: 60.937 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion:129 Resp: 97360
Ion Ratio Lower Upper
129 100
127 76.8 38.7 116.1





#61
1,2-Dibromoethane
Concen: 56.707 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:107 Resp: 75264
Ion Ratio Lower Upper
107 100
109 94.7 75.4 113.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#62
4-Bromofluorobenzene
Concen: 48.565 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion: 95 Resp: 107492
Ion Ratio Lower Upper
95 100
174 88.5 0.0 177.0
176 85.3 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 1586
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:117 Resp: 249980
Ion Ratio Lower Upper
117 100
82 53.9 43.0 64.4
119 32.2 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#64
Tetrachloroethene
Concen: 49.135 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion:164 Resp: 78624
Ion Ratio Lower Upper
164 100
166 125.7 100.4 150.6
129 87.0 71.4 107.0
131 86.8 70.8 106.2





#65
Chlorobenzene
Concen: 52.940 ug/l
RT: 11.514 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:112 Resp: 267840
Ion Ratio Lower Upper
112 100
114 30.7 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 57.251 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion:131 Resp: 105820
Ion Ratio Lower Upper
131 100
133 93.9 47.9 143.6
119 66.3 30.8 92.3





#67
Ethyl Benzene
Concen: 54.421 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 491700
Ion Ratio Lower Upper
91 100
106 31.5 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#68
m/p-Xylenes
Concen: 113.078 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion:106 Resp: 407016
Ion Ratio Lower Upper
106 100
91 194.9 166.6 249.8





#69
o-Xylene
Concen: 58.494 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:106 Resp: 194540
Ion Ratio Lower Upper
106 100
91 205.0 110.3 331.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#70
Styrene
Concen: 61.253 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion:104 Resp: 344857
Ion Ratio Lower Upper
104 100
78 45.3 42.4 63.6
103 53.6 44.4 66.6





#71
Bromoform
Concen: 55.458 ug/l
RT: 12.209 min Scan# 1704
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:173 Resp: 6958
Ion Ratio Lower Upper
173 100
175 47.6 24.6 74.0
254 0.1 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion:152 Resp: 153750
Ion Ratio Lower Upper
152 100
115 57.1 28.2 84.7
150 177.0 0.0 347.6





#73
Isopropylbenzene
Concen: 49.913 ug/l
RT: 12.331 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:105 Resp: 52632
Ion Ratio Lower Upper
105 100
120 26.7 13.4 40.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#74
N-ethyl acetate
Concen: 50.802 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion: 43 Resp: 127866
Ion Ratio Lower Upper
43 100
70 48.0 28.7 43.1#
55 26.2 17.3 25.9#
61 25.7 17.0 25.6#





#75
1,1,2,2-Tetrachloroethane
Concen: 45.469 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 83 Resp: 106590
Ion Ratio Lower Upper
83 100
131 10.0 5.6 16.8
85 65.9 32.3 96.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#76
1,2,3-Trichloropropane
Concen: 45.287 ug/l m
RT: 12.635 min Scan# 1774
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13



Tgt Ion: 75 Resp: 80431
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 49.700 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:156 Resp: 121433
Ion Ratio Lower Upper
156 100
77 177.1 95.5 286.4
158 97.0 48.0 144.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#78
n-propylbenzene
Concen: 50.090 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13

Tgt Ion: 91 Resp: 648876
Ion Ratio Lower Upper
91 100
120 23.1 11.3 33.9









#79
2-Chlorotoluene
Concen: 51.553 ug/l
RT: 12.757 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 361719
Ion Ratio Lower Upper
91 100
126 35.6 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#80
1,3,5-Trimethylbenzene
Concen: 54.025 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion:105 Resp: 459962
Ion Ratio Lower Upper
105 100
120 50.6 24.4 73.4





#81
trans-1,4-Dichloro-2-butene
Concen: 44.671 ug/l
RT: 12.379 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 75 Resp: 36272
Ion Ratio Lower Upper
75 100
53 85.1 85.3 127.9
89 46.4 37.0 55.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#82
4-Chlorotoluene
Concen: 53.893 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion: 91 Resp: 399480
Ion Ratio Lower Upper
91 100
126 35.6 17.0 51.0





#83
tert-Butylbenzene
Concen: 53.404 ug/l
RT: 13.074 min Scan# 1846
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:119 Resp: 408900
Ion Ratio Lower Upper
119 100
91 63.0 34.8 104.4
134 27.5 13.3 39.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#84
1,2,4-Trimethylbenzene
Concen: 55.078 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion:105 Resp: 464219
Ion Ratio Lower Upper
105 100
120 46.9 22.3 66.9





#85
 sec-Butylbenzene
 Concen: 53.221 ug/l
 RT: 13.257 min Scan# 1876
 Delta R.T. 0.006 min
 Lab File: VY008883.D
 Acq: 19 May 2022 21:13




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion:105 Resp: 635772
 Ion Ratio Lower Upper
 105 100
 134 20.3 10.2 30.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 05/24/2022
 Supervised By :Mahesh Dadoda 05/24/2022




#86
 p-Isopropyltoluene
 Concen: 54.623 ug/l
 RT: 13.373 min Scan# 1895
 Delta R.T. 0.006 min
 Lab File: VY008883.D
 Acq: 19 May 2022 21:13




Tgt Ion:119 Resp: 532059
 Ion Ratio Lower Upper
 119 100
 134 27.0 13.2 39.5
 91 22.7 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 53.575 ug/l
RT: 13.367 min Scan# 1191
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 275040
Ion Ratio Lower Upper
146 100
111 38.3 19.5 58.5
148 63.3 31.6 95.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#88
1,4-Dichlorobenzene
Concen: 53.828 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion:146 Resp: 273340
Ion Ratio Lower Upper
146 100
111 37.3 19.3 57.8
148 64.3 32.6 97.8





#89
n-Butylbenzene
Concen: 51.508 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 476224
Ion Ratio Lower Upper
91 100
92 52.9 26.3 78.9
134 25.9 13.0 38.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#90
Hexachloroethane
Concen: 53.166 ug/l
RT: 13.958 min Scan# 1991
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13


Tgt Ion:117 Resp: 97912
Ion Ratio Lower Upper
117 100
201 71.0 43.0 128.8





#91
1,2-Dichlorobenzene
Concen: 54.616 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 245200
Ion Ratio Lower Upper
146 100
111 39.7 20.3 60.9
148 63.1 32.3 96.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 39.304 ug/l
RT: 14.361 min Scan# 2057
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion: 75 Resp: 15754
Ion Ratio Lower Upper
75 100
155 96.8 38.0 113.9
157 120.6 47.8 143.3





#93
1,2,4-Trichlorobenzene
Concen: 52.276 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:180 Resp: 128162
Ion Ratio Lower Upper
180 100
182 93.4 47.3 141.9
145 29.4 14.6 43.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022




#94
Hexachlorobutadiene
Concen: 48.698 ug/l
RT: 15.117 min Scan# 2181
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Tgt Ion:225 Resp: 68701
Ion Ratio Lower Upper
225 100
223 63.1 30.8 92.3
227 64.3 31.6 94.7





#95
Naphthalene
Concen: 48.770 ug/l
RT: 15.239 min Scan# 21
Delta R.T. -0.000 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:128 Resp: 267292
Ion Ratio Lower Upper
128 100
127 13.5 11.0 16.4
129 11.0 8.4 12.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022



#96
1,2,3-Trichlorobenzene
Concen: 52.729 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. 0.006 min
Lab File: VY008883.D
Acq: 19 May 2022 21:13



Tgt Ion:180 Resp: 113561
Ion Ratio Lower Upper
180 100
182 94.5 47.3 141.9
145 30.5 15.3 45.9

