



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052224\
Data File : VY018342.D
Acq On : 22 May 2024 08:42
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052224S.M
Title : SW846 8260
Last Update : Thu May 23 01:40:36 2024



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	13831	PASS
75	95	30	60	50.4	35699	PASS
95	95	100	100	100.0	70765	PASS
96	95	5	9	6.2	4403	PASS
173	174	0.00	2	0.9	511	PASS
174	95	50	100	81.4	57603	PASS
175	174	5	9	7.0	4037	PASS
176	174	95	101	95.7	55125	PASS
177	176	5	9	6.7	3712	PASS