

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052225\
 Data File : VY022384.D
 Acq On : 22 May 2025 10:38
 Operator : SY/MD
 Sample : Q2087-03RE
 Misc : 5.38g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-08-05202025RE

Quant Time: May 23 03:33:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:42:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	32360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	45412	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	27449	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.353	152	7819	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	11107	31.405	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	62.820%
35) Dibromofluoromethane	7.640	113	10942	40.370	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	80.740%
50) Toluene-d8	10.115	98	49381	44.484	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.960%
62) 4-Bromofluorobenzene	12.414	95	9237	26.252	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	52.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052225\
Data File : VY022384.D
Acq On : 22 May 2025 10:38
Operator : SY/MD
Sample : Q2087-03RE
Misc : 5.38g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-08-05202025RE

Quant Time: May 23 03:33:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:42:09 2025
Response via : Initial Calibration

