

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052225\
 Data File : VY022388.D
 Acq On : 22 May 2025 12:11
 Operator : SY/MD
 Sample : Q2104-01RE
 Misc : 5.21g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-052125RE

Quant Time: May 23 03:35:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:42:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.720	168	7444	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	8.622	114	11850	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	9162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	2096	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	3455	42.468	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	84.940%
35) Dibromofluoromethane	7.646	113	2050	28.984	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	57.960%
50) Toluene-d8	10.109	98	13844	47.792	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.580%
62) 4-Bromofluorobenzene	12.408	95	2941	32.032	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	64.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052225\
Data File : VY022388.D
Acq On : 22 May 2025 12:11
Operator : SY/MD
Sample : Q2104-01RE
Misc : 5.21g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-052125RE

Quant Time: May 23 03:35:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:42:09 2025
Response via : Initial Calibration

