

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052725\
 Data File : VY022436.D
 Acq On : 27 May 2025 16:01
 Operator : SY/MD
 Sample : Q2130-01
 Misc : 5.21g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3

Quant Time: May 28 08:01:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052725S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 28 10:34:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	301920	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	554808	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	448596	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	169776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	167276	52.600	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.200%
35) Dibromofluoromethane	7.634	113	167053	51.415	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.820%
50) Toluene-d8	10.103	98	666628	50.330	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.660%
62) 4-Bromofluorobenzene	12.408	95	172667	43.097	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052725\
Data File : VY022436.D
Acq On : 27 May 2025 16:01
Operator : SY/MD
Sample : Q2130-01
Misc : 5.21g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3

Quant Time: May 28 08:01:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052725S.M
Quant Title : SW846 8260
QLast Update : Wed May 28 10:34:06 2025
Response via : Initial Calibration

