

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052925\
 Data File : VY022454.D
 Acq On : 29 May 2025 08:56
 Operator : SY/MD
 Sample : VY0529SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0529SBL01

Quant Time: May 30 05:29:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052725S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 28 10:34:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	320837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.621	114	584950	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	455837	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	159461	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	180255	53.339	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.680%
35) Dibromofluoromethane	7.646	113	173894	50.762	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.520%
50) Toluene-d8	10.115	98	699660	50.101	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.200%
62) 4-Bromofluorobenzene	12.413	95	172737	40.893	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052925\
Data File : VY022454.D
Acq On : 29 May 2025 08:56
Operator : SY/MD
Sample : VY0529SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBL01

Quant Time: May 30 05:29:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052725S.M
Quant Title : SW846 8260
QLast Update : Wed May 28 10:34:06 2025
Response via : Initial Calibration

