

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053024\
 Data File : VY018410.D
 Acq On : 30 May 2024 13:29
 Operator : SY/MD
 Sample : P2610-03
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-05242024

Quant Time: May 31 01:01:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052224S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 23 01:40:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	93714	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	189816	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	170100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	60291	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	57840	65.145	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 130.300%	
35) Dibromofluoromethane	7.710	113	59001	53.302	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.600%	
50) Toluene-d8	10.179	98	227561	52.729	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 105.460%	
62) 4-Bromofluorobenzene	12.483	95	67137	50.346	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 100.700%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053024\
Data File : VY018410.D
Acq On : 30 May 2024 13:29
Operator : SY/MD
Sample : P2610-03
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-02-05242024

Quant Time: May 31 01:01:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052224S.M
Quant Title : SW846 8260
QLast Update : Thu May 23 01:40:36 2024
Response via : Initial Calibration

