

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053025\
 Data File : VY022487.D
 Acq On : 30 May 2025 18:57
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 31 02:32:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052725S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 28 10:34:06 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	73	0.00
2 T	Dichlorodifluoromethane	0.477	0.416	12.8	71	0.00
3 P	Chloromethane	1.173	1.173	0.0	71	0.00
4 C	Vinyl Chloride	1.425	1.623	-13.9#	81	0.00
5 T	Bromomethane	1.295	1.328	-2.5	76	0.00
6 T	Chloroethane	0.931	1.147	-23.2	82	0.00
7 T	Trichlorofluoromethane	1.203	1.350	-12.2	82	0.00
8 T	Diethyl Ether	0.277	0.305	-10.1	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.528	0.534	-1.1	74	0.00
10 T	Methyl Iodide	0.595	0.636	-6.9	71	-0.01
11 T	Tert butyl alcohol	0.037	0.042	-13.5	85	-0.02
12 CM	1,1-Dichloroethene	0.524	0.533	-1.7#	76	0.00
13 T	Acrolein	0.056	0.040	28.6#	51	-0.01
14 T	Allyl chloride	0.788	0.839	-6.5	77	0.00
15 T	Acrylonitrile	0.112	0.129	-15.2	84	-0.01
16 T	Acetone	0.087	0.110	-26.4#	94	0.00
17 T	Carbon Disulfide	1.672	1.679	-0.4	74	0.00
18 T	Methyl Acetate	0.324	0.417	-28.7#	98	-0.01
19 T	Methyl tert-butyl Ether	1.391	1.512	-8.7	79	0.00
20 T	Methylene Chloride	0.611	0.723	-18.3	96	0.00
21 T	trans-1,2-Dichloroethene	0.572	0.593	-3.7	77	0.00
22 T	Diisopropyl ether	1.739	1.884	-8.3	79	0.00
23 T	Vinyl Acetate	0.999	1.042	-4.3	75	-0.01
24 P	1,1-Dichloroethane	1.028	1.125	-9.4	79	0.00
25 T	2-Butanone	0.144	0.166	-15.3	85	-0.01
26 T	2,2-Dichloropropane	0.931	0.909	2.4	72	0.00
27 T	cis-1,2-Dichloroethene	0.662	0.712	-7.6	79	0.00
28 T	Bromochloromethane	0.429	0.499	-16.3	83	0.00
29 T	Tetrahydrofuran	0.096	0.109	-13.5	83	0.00
30 C	Chloroform	1.013	1.128	-11.4#	80	0.00
31 T	Cyclohexane	1.013	0.956	5.6	74	0.00
32 T	1,1,1-Trichloroethane	0.919	0.971	-5.7	77	0.00
33 S	1,2-Dichloroethane-d4	0.527	0.611	-15.9	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	75	0.00
35 S	Dibromofluoromethane	0.293	0.326	-11.3	85	0.00
36 T	1,1-Dichloropropene	0.465	0.472	-1.5	77	0.00
37 T	Ethyl Acetate	0.197	0.207	-5.1	78	0.00
38 T	Carbon Tetrachloride	0.484	0.487	-0.6	76	0.00
39 T	Methylcyclohexane	0.618	0.599	3.1	73	0.00
40 TM	Benzene	1.383	1.471	-6.4	79	0.00
41 T	Methacrylonitrile	0.120	0.127	-5.8	81	-0.02
42 TM	1,2-Dichloroethane	0.365	0.395	-8.2	80	0.00
43 T	Isopropyl Acetate	0.432	0.463	-7.2	80	0.00
44 TM	Trichloroethene	0.342	0.348	-1.8	76	0.00
45 C	1,2-Dichloropropane	0.325	0.348	-7.1#	81	0.00
46 T	Dibromomethane	0.181	0.195	-7.7	80	0.00
47 T	Bromodichloromethane	0.467	0.496	-6.2	80	0.00
48 T	Methyl methacrylate	0.200	0.219	-9.5	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053025\
 Data File : VY022487.D
 Acq On : 30 May 2025 18:57
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 31 02:32:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052725S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 28 10:34:06 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	89	0.00
50 S	Toluene-d8	1.194	1.297	-8.6	83	0.00
51 T	4-Methyl-2-Pentanone	0.213	0.238	-11.7	82	0.00
52 CM	Toluene	0.864	0.917	-6.1#	79	0.00
53 T	t-1,3-Dichloropropene	0.435	0.452	-3.9	78	0.00
54 T	cis-1,3-Dichloropropene	0.511	0.532	-4.1	79	0.00
55 T	1,1,2-Trichloroethane	0.229	0.248	-8.3	81	0.00
56 T	Ethyl methacrylate	0.329	0.355	-7.9	78	0.00
57 T	1,3-Dichloropropane	0.404	0.445	-10.1	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.164	0.184	-12.2	83	0.00
59 T	2-Hexanone	0.140	0.159	-13.6	82	0.00
60 T	Dibromochloromethane	0.293	0.312	-6.5	78	0.00
61 T	1,2-Dibromoethane	0.210	0.231	-10.0	81	0.00
62 S	4-Bromofluorobenzene	0.361	0.390	-8.0	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	74	0.00
64 T	Tetrachloroethene	0.426	0.458	-7.5	79	0.00
65 PM	Chlorobenzene	1.066	1.128	-5.8	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.388	-7.5	78	0.00
67 C	Ethyl Benzene	1.980	2.086	-5.4#	78	0.00
68 T	m/p-Xylenes	0.751	0.782	-4.1	77	0.00
69 T	o-Xylene	0.705	0.737	-4.5	77	0.00
70 T	Styrene	1.175	1.263	-7.5	79	0.00
71 P	Bromoform	0.189	0.199	-5.3	77	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
73 T	Isopropylbenzene	3.935	4.013	-2.0	77	0.00
74 T	N-amyl acetate	0.963	1.004	-4.3	74	0.00
75 P	1,1,2,2-Tetrachloroethane	0.620	0.657	-6.0	79	0.00
76 T	1,2,3-Trichloropropane	0.501	0.487	2.8	73	0.00
77 T	Bromobenzene	0.847	0.866	-2.2	75	0.00
78 T	n-propylbenzene	4.723	4.859	-2.9	76	0.00
79 T	2-Chlorotoluene	2.556	2.642	-3.4	78	0.00
80 T	1,3,5-Trimethylbenzene	3.075	3.173	-3.2	76	0.00
81 T	trans-1,4-Dichloro-2-butene	0.221	0.215	2.7	74	0.00
82 T	4-Chlorotoluene	2.633	2.688	-2.1	76	0.00
83 T	tert-Butylbenzene	2.778	2.852	-2.7	77	0.00
84 T	1,2,4-Trimethylbenzene	3.094	3.190	-3.1	76	0.00
85 T	sec-Butylbenzene	4.120	4.264	-3.5	76	0.00
86 T	p-Isopropyltoluene	3.440	3.521	-2.4	75	0.00
87 T	1,3-Dichlorobenzene	1.704	1.734	-1.8	76	0.00
88 T	1,4-Dichlorobenzene	1.629	1.674	-2.8	76	0.00
89 T	n-Butylbenzene	3.226	3.295	-2.1	73	0.00
90 T	Hexachloroethane	0.685	0.701	-2.3	75	0.00
91 T	1,2-Dichlorobenzene	1.423	1.488	-4.6	78	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.096	0.101	-5.2	80	0.00
93 T	1,2,4-Trichlorobenzene	0.789	0.785	0.5	70	0.00
94 T	Hexachlorobutadiene	0.414	0.411	0.7	72	0.00
95 T	Naphthalene	1.459	1.583	-8.5	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053025\
Data File : VY022487.D
Acq On : 30 May 2025 18:57
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: May 31 02:32:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052725S.M
Quant Title : SW846 8260
QLast Update : Wed May 28 10:34:06 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.647	0.668	-3.2	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6