

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060324\
 Data File : VY018435.D
 Acq On : 03 Jun 2024 12:34
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jun 04 05:28:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 05:22:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	92688	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	183124	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	158367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	56616	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	52575	28.883	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	57.760%
35) Dibromofluoromethane	7.710	113	55526	23.912	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	47.820%#
50) Toluene-d8	10.173	98	217042	23.494	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	46.980%#
62) 4-Bromofluorobenzene	12.477	95	61984	21.550	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	43.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060324\
Data File : VY018435.D
Acq On : 03 Jun 2024 12:34
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jun 04 05:28:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 05:22:03 2024
Response via : Initial Calibration

