

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060425\
 Data File : VY022545.D
 Acq On : 04 Jun 2025 13:03
 Operator : SY/MD
 Sample : Q2182-01RE
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-06022025RE

Quant Time: Jun 05 04:14:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	33307	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	52550	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	35151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	9867	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	16642	44.674	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.340%
35) Dibromofluoromethane	7.634	113	14848	47.338	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.680%
50) Toluene-d8	10.109	98	57838	45.636	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.280%
62) 4-Bromofluorobenzene	12.401	95	11930	31.593	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	63.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060425\
Data File : VY022545.D
Acq On : 04 Jun 2025 13:03
Operator : SY/MD
Sample : Q2182-01RE
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-06022025RE

Quant Time: Jun 05 04:14:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 03 03:22:04 2025
Response via : Initial Calibration

