

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060425\
 Data File : VY022557.D
 Acq On : 04 Jun 2025 17:45
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jun 05 04:19:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	370158	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.610	114	676276	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	551141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	208389	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	205626	49.668	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.340%
35) Dibromofluoromethane	7.628	113	204231	50.595	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.200%
50) Toluene-d8	10.103	98	825747	50.627	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.260%
62) 4-Bromofluorobenzene	12.402	95	209589	43.129	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060425\
 Data File : VY022557.D
 Acq On : 04 Jun 2025 17:45
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
ClientSampleId :
 VIBLK

Quant Time: Jun 05 04:19:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

