

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060525\
 Data File : VY022561.D
 Acq On : 05 Jun 2025 09:11
 Operator : SY/MD
 Sample : VY0605SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0605SBL01

Quant Time: Jun 06 01:19:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	260317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	481555	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	383268	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	134444	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.060	65	150403	51.658	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.320%
35) Dibromofluoromethane	7.634	113	144501	50.273	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.540%
50) Toluene-d8	10.103	98	576294	49.621	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.240%
62) 4-Bromofluorobenzene	12.401	95	140831	40.698	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060525\
Data File : VY022561.D
Acq On : 05 Jun 2025 09:11
Operator : SY/MD
Sample : VY0605SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0605SBL01

Quant Time: Jun 06 01:19:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 03 03:22:04 2025
Response via : Initial Calibration

