

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060525\
 Data File : VY022577.D
 Acq On : 05 Jun 2025 16:19
 Operator : SY/MD
 Sample : Q2207-11
 Misc : 5.92g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BU-703-VOC-02

Quant Time: Jun 06 01:26:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	505004	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.610	114	940265	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	784646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	300101	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	302154	53.496	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.000%
35) Dibromofluoromethane	7.628	113	287515	51.230	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.460%
50) Toluene-d8	10.103	98	1167445	51.481	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.960%
62) 4-Bromofluorobenzene	12.402	95	304388	45.050	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060525\
Data File : VY022577.D
Acq On : 05 Jun 2025 16:19
Operator : SY/MD
Sample : Q2207-11
Misc : 5.92g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BU-703-VOC-02

Quant Time: Jun 06 01:26:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 03 03:22:04 2025
Response via : Initial Calibration

