



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060525\
Data File : VY022559.D
Acq On : 05 Jun 2025 08:08
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Title : SW846 8260
Last Update : Tue Jun 03 03:22:04 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	25.2	21955	PASS
75	95	30	60	58.6	51101	PASS
95	95	100	100	100.0	87144	PASS
96	95	5	9	6.6	5734	PASS
173	174	0.00	2	0.9	652	PASS
174	95	50	100	83.3	72632	PASS
175	174	5	9	7.9	5704	PASS
176	174	95	101	97.1	70496	PASS
177	176	5	9	6.4	4509	PASS