

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060724\
 Data File : VY018527.D
 Acq On : 07 Jun 2024 15:14
 Operator : SY/MD
 Sample : P2707-06
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SW-04

Quant Time: Jun 07 23:13:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	1690	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	1190	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	237	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.124	65	1453	89.900	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	179.800%#
35) Dibromofluoromethane	7.698	113	969	45.216	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.440%
50) Toluene-d8	10.179	98	2745	32.197	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	64.400%
62) 4-Bromofluorobenzene	12.489	95	539	20.305	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	40.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060724\
Data File : VY018527.D
Acq On : 07 Jun 2024 15:14
Operator : SY/MD
Sample : P2707-06
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SW-04

Quant Time: Jun 07 23:13:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration

