

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060724\
 Data File : VY018528.D
 Acq On : 07 Jun 2024 15:39
 Operator : SY/MD
 Sample : P2707-04
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SW-02

Quant Time: Jun 07 23:14:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	1145	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	2366	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	1464	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	215	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	1673	74.402	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	148.800%
35) Dibromofluoromethane	7.704	113	1256	41.863	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.720%
50) Toluene-d8	10.179	98	3431	28.746	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	57.500%#
62) 4-Bromofluorobenzene	12.483	95	544	14.638	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	29.280%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060724\
Data File : VY018528.D
Acq On : 07 Jun 2024 15:39
Operator : SY/MD
Sample : P2707-04
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SW-02

Quant Time: Jun 07 23:14:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration

