

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060724\
 Data File : VY018519.D
 Acq On : 07 Jun 2024 10:59
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 07 23:08:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	0.794	0.672	15.4	77	0.00
3 P	Chloromethane	1.481	1.568	-5.9	91	0.00
4 C	Vinyl Chloride	1.771	1.840	-3.9#	89	0.00
5 T	Bromomethane	1.182	1.210	-2.4	90	0.00
6 T	Chloroethane	1.118	1.225	-9.6	90	0.00
7 T	Trichlorofluoromethane	1.833	1.824	0.5	85	0.00
8 T	Diethyl Ether	0.551	0.520	5.6	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	1.067	1.002	6.1	82	0.00
10 T	Methyl Iodide	1.048	1.041	0.7	79	0.00
11 T	Tert butyl alcohol	0.171	0.072	57.9#	54	0.00
12 CM	1,1-Dichloroethene	1.049	0.992	5.4#	81	0.00
13 T	Acrolein	0.101	0.096	5.0	56	0.00
14 T	Allyl chloride	1.536	1.549	-0.8	81	0.00
15 T	Acrylonitrile	0.252	0.244	3.2	76	0.00
16 T	Acetone	0.218	0.263	-20.6	88	0.00
17 T	Carbon Disulfide	3.077	3.017	1.9	85	0.00
18 T	Methyl Acetate	0.545	0.514	5.7	80	0.00
19 T	Methyl tert-butyl Ether	2.528	2.414	4.5	76	-0.01
20 T	Methylene Chloride	1.348	1.213	10.0	86	0.00
21 T	trans-1,2-Dichloroethene	1.119	1.096	2.1	80	0.00
22 T	Diisopropyl ether	3.128	3.374	-7.9	85	0.00
23 T	Vinyl Acetate	2.036	2.186	-7.4	82	0.00
24 P	1,1-Dichloroethane	1.951	1.960	-0.5	82	0.00
25 T	2-Butanone	0.334	0.355	-6.3	77	0.00
26 T	2,2-Dichloropropane	1.715	1.604	6.5	78	0.00
27 T	cis-1,2-Dichloroethene	1.302	1.261	3.1	79	0.00
28 T	Bromochloromethane	0.898	0.985	-9.7	83	0.00
29 T	Tetrahydrofuran	0.217	0.218	-0.5	79	0.00
30 C	Chloroform	1.957	1.958	-0.1#	81	0.00
31 T	Cyclohexane	1.831	1.843	-0.7	85	0.00
32 T	1,1,1-Trichloroethane	1.714	1.642	4.2	79	0.00
33 S	1,2-Dichloroethane-d4	0.982	1.073	-9.3	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	74	0.00
35 S	Dibromofluoromethane	0.634	0.672	-6.0	82	0.00
36 T	1,1-Dichloropropene	0.967	0.971	-0.4	82	0.00
37 T	Ethyl Acetate	0.483	0.497	-2.9	82	0.00
38 T	Carbon Tetrachloride	0.969	0.911	6.0	80	0.00
39 T	Methylcyclohexane	1.307	1.278	2.2	83	0.00
40 TM	Benzene	2.897	2.914	-0.6	81	0.00
41 T	Methacrylonitrile	0.235	0.246	-4.7	81	0.00
42 TM	1,2-Dichloroethane	0.704	0.726	-3.1	82	0.00
43 T	Isopropyl Acetate	0.903	0.910	-0.8	76	0.00
44 TM	Trichloroethene	0.718	0.676	5.8	78	0.00
45 C	1,2-Dichloropropane	0.662	0.692	-4.5#	83	0.00
46 T	Dibromomethane	0.380	0.386	-1.6	82	0.00
47 T	Bromodichloromethane	0.912	0.925	-1.4	81	0.00
48 T	Methyl methacrylate	0.415	0.420	-1.2	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060724\
 Data File : VY018519.D
 Acq On : 07 Jun 2024 10:59
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 07 23:08:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.004	20.0	74	-0.01
50 S	Toluene-d8	2.522	2.668	-5.8	82	0.00
51 T	4-Methyl-2-Pentanone	0.469	0.478	-1.9	75	0.00
52 CM	Toluene	1.782	1.847	-3.6#	81	0.00
53 T	t-1,3-Dichloropropene	0.837	0.870	-3.9	80	0.00
54 T	cis-1,3-Dichloropropene	1.031	1.040	-0.9	80	0.00
55 T	1,1,2-Trichloroethane	0.494	0.501	-1.4	77	0.00
56 T	Ethyl methacrylate	0.695	0.715	-2.9	76	0.00
57 T	1,3-Dichloropropane	0.839	0.896	-6.8	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.373	0.392	-5.1	74	0.00
59 T	2-Hexanone	0.324	0.360	-11.1	79	0.00
60 T	Dibromochloromethane	0.604	0.583	3.5	76	0.00
61 T	1,2-Dibromoethane	0.465	0.456	1.9	76	0.00
62 S	4-Bromofluorobenzene	0.785	0.842	-7.3	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.736	0.661	10.2	82	0.00
65 PM	Chlorobenzene	2.144	2.015	6.0	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.693	0.641	7.5	80	0.00
67 C	Ethyl Benzene	3.903	3.786	3.0#	82	0.00
68 T	m/p-Xylenes	1.470	1.429	2.8	82	0.00
69 T	o-Xylene	1.415	1.312	7.3	80	0.00
70 T	Styrene	2.311	2.267	1.9	81	0.00
71 P	Bromoform	0.365	0.330	9.6	76	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	5.160	4.805	6.9	83	0.00
74 T	N-amyl acetate	1.256	1.221	2.8	83	0.00
75 P	1,1,2,2-Tetrachloroethane	0.908	0.821	9.6	78	0.00
76 T	1,2,3-Trichloropropane	0.620	0.558	10.0	83	0.00
77 T	Bromobenzene	1.096	0.959	12.5	78	0.00
78 T	n-propylbenzene	6.193	5.921	4.4	87	0.00
79 T	2-Chlorotoluene	3.367	3.210	4.7	87	0.00
80 T	1,3,5-Trimethylbenzene	4.019	3.794	5.6	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.295	0.252	14.6	72	0.00
82 T	4-Chlorotoluene	3.388	3.248	4.1	91	0.00
83 T	tert-Butylbenzene	3.716	3.440	7.4	82	0.00
84 T	1,2,4-Trimethylbenzene	3.934	3.825	2.8	91	0.00
85 T	sec-Butylbenzene	5.473	5.267	3.8	90	0.00
86 T	p-Isopropyltoluene	4.417	4.298	2.7	90	0.00
87 T	1,3-Dichlorobenzene	2.161	2.038	5.7	85	0.00
88 T	1,4-Dichlorobenzene	2.123	1.966	7.4	85	0.00
89 T	n-Butylbenzene	4.197	4.209	-0.3	96	0.00
90 T	Hexachloroethane	0.828	0.770	7.0	87	0.00
91 T	1,2-Dichlorobenzene	1.880	1.747	7.1	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.138	0.116	15.9	83	0.00
93 T	1,2,4-Trichlorobenzene	1.017	0.881	13.4	81	0.00
94 T	Hexachlorobutadiene	0.540	0.450	16.7	82	0.00
95 T	Naphthalene	2.043	1.739	14.9	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060724\
Data File : VY018519.D
Acq On : 07 Jun 2024 10:59
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jun 07 23:08:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.860	0.726	15.6	78	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6