



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060724\
Data File : VY018518.D
Acq On : 07 Jun 2024 09:08
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Title : SW846 8260
Last Update : Tue Jun 04 07:11:56 2024



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.4	5908	PASS
75	95	30	60	58.2	13520	PASS
95	95	100	100	100.0	23216	PASS
96	95	5	9	6.3	1473	PASS
173	174	0.00	2	1.4	260	PASS
174	95	50	100	79.1	18357	PASS
175	174	5	9	6.7	1229	PASS
176	174	95	101	96.6	17729	PASS
177	176	5	9	7.2	1284	PASS