

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060925\
 Data File : VY022622.D
 Acq On : 09 Jun 2025 12:02
 Operator : SY/MD
 Sample : Q2240-07
 Misc : 4.28g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-2-VOC

Quant Time: Jun 10 01:27:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	299798	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	551339	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	439182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	156038	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	153610	45.812	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.620%
35) Dibromofluoromethane	7.640	113	162973	49.523	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.040%
50) Toluene-d8	10.109	98	671309	50.486	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.980%
62) 4-Bromofluorobenzene	12.408	95	161567	40.781	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.560%
Target Compounds						
16) Acetone	3.873	43	7742	11.369	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060925\
Data File : VY022622.D
Acq On : 09 Jun 2025 12:02
Operator : SY/MD
Sample : Q2240-07
Misc : 4.28g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-2-VOC

Quant Time: Jun 10 01:27:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 03 03:22:04 2025
Response via : Initial Calibration

