

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060925\
 Data File : VY022626.D
 Acq On : 09 Jun 2025 13:36
 Operator : SY/MD
 Sample : Q2244-02
 Misc : 5.96g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP03-MHC-VOC

Quant Time: Jun 10 01:28:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	285760	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	530976	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	426137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	152100	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	153629	48.068	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.140%
35) Dibromofluoromethane	7.640	113	159802	50.422	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.840%
50) Toluene-d8	10.109	98	642201	50.149	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.300%
62) 4-Bromofluorobenzene	12.408	95	155434	40.737	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060925\
Data File : VY022626.D
Acq On : 09 Jun 2025 13:36
Operator : SY/MD
Sample : Q2244-02
Misc : 5.96g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP03-MHC-VOC

Quant Time: Jun 10 01:28:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 03 03:22:04 2025
Response via : Initial Calibration

