

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061024\
 Data File : VY018548.D
 Acq On : 10 Jun 2024 14:12
 Operator : SY/MD
 Sample : P2786-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-02-060724

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/11/2024
 Supervised By :Semsettin Yesilyurt 06/11/2024

Quant Time: Jun 11 01:55:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	377	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	1089m	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	575m	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	98	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	985	133.042	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	266.080%#
35) Dibromofluoromethane	7.655	113	61	4.417	ug/l	-0.05
Spiked Amount	50.000	Range	54 - 147	Recovery	=	8.840%#
50) Toluene-d8	10.179	98	1790	32.583	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	65.160%
62) 4-Bromofluorobenzene	12.483	95	143	8.360	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	16.720%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061024\
Data File : VY018548.D
Acq On : 10 Jun 2024 14:12
Operator : SY/MD
Sample : P2786-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-02-060724

Quant Time: Jun 11 01:55:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/11/2024
Supervised By :Semsettin Yesilyurt 06/11/2024

